

## ANALYTICAL REPORT


Job Number: 280-111519-2

Job Description: Ravenna, OH

For:

Cardno GS, Inc  
2496 Old Ivy Road  
Suite 300  
Charlottesville, VA 22903

Attention: Mr. Peter Chapman



Approved for release.  
Patrick J McEntee  
Manager of Project Management  
10/4/2018 3:15 PM

---

Patrick J McEntee, Manager of Project Management  
4955 Yarrow Street, Arvada, CO, 80002  
(303)736-0107  
patrick.mcatee@testamericainc.com  
10/04/2018

The test results in this report relate only to the samples in this report and meet all requirements of NELAC, with any exceptions noted. Pursuant to NELAP, this report shall not be reproduced except in full, without the written approval of the laboratory. All questions regarding this report should be directed to the TestAmerica Denver Project Manager.

The Lab Certification ID# is 4025.

Reporting limits are adjusted for sample size used, dilutions and moisture content if applicable.

# Table of Contents

Cover Title Page .....	1
Data Summaries .....	4
Definitions .....	4
Case Narrative .....	5
Detection Summary .....	7
Client Sample Results .....	8
Default Detection Limits .....	9
QC Sample Results .....	10
QC Association .....	12
Chronicle .....	13
Certification Summary .....	14
Method Summary .....	15
Sample Summary .....	16
Manual Integration Summary .....	17
Reagent Traceability .....	18
COAs .....	19
Inorganic Sample Data .....	31
General Chemistry Data .....	31
Gen Chem Cover Page .....	32
Gen Chem Sample Data .....	33
Gen Chem QC Data .....	35
Gen Chem ICV/CCV .....	35
Gen Chem Blanks .....	37
Gen Chem MS/MSD/PDS .....	38
Gen Chem Duplicates .....	40
Gen Chem LCS/LCSD .....	41

# Table of Contents

Gen Chem MDL .....	44
Gen Chem Analysis Run Log .....	46
Gen Chem Prep Data .....	50
Gen Chem Raw Data .....	54
<b>Shipping and Receiving Documents .....</b>	<b>106</b>
Client Chain of Custody .....	107
Sample Receipt Checklist .....	114

# Definitions/Glossary

Client: Cardno GS, Inc  
Project/Site: Ravenna, OH

TestAmerica Job ID: 280-111519-2

## Qualifiers

### General Chemistry

Qualifier	Qualifier Description
H	Sample was prepped or analyzed beyond the specified holding time
J	Estimated: The analyte was positively identified; the quantitation is an estimation
U	Undetected at the Limit of Detection.

## Glossary

Abbreviation	These commonly used abbreviations may or may not be present in this report.
¤	Listed under the "D" column to designate that the result is reported on a dry weight basis
%R	Percent Recovery
CFL	Contains Free Liquid
CNF	Contains No Free Liquid
DER	Duplicate Error Ratio (normalized absolute difference)
Dil Fac	Dilution Factor
DL	Detection Limit (DoD/DOE)
DL, RA, RE, IN	Indicates a Dilution, Re-analysis, Re-extraction, or additional Initial metals/anion analysis of the sample
DLC	Decision Level Concentration (Radiochemistry)
EDL	Estimated Detection Limit (Dioxin)
LOD	Limit of Detection (DoD/DOE)
LOQ	Limit of Quantitation (DoD/DOE)
MDA	Minimum Detectable Activity (Radiochemistry)
MDC	Minimum Detectable Concentration (Radiochemistry)
MDL	Method Detection Limit
ML	Minimum Level (Dioxin)
NC	Not Calculated
ND	Not Detected at the reporting limit (or MDL or EDL if shown)
PQL	Practical Quantitation Limit
QC	Quality Control
RER	Relative Error Ratio (Radiochemistry)
RL	Reporting Limit or Requested Limit (Radiochemistry)
RPD	Relative Percent Difference, a measure of the relative difference between two points
TEF	Toxicity Equivalent Factor (Dioxin)
TEQ	Toxicity Equivalent Quotient (Dioxin)

## CASE NARRATIVE

**Client: Cardno GS, Inc**

**Project: Ravenna, OH**

**Report Number: 280-111519-2**

With the exceptions noted as flags or footnotes, standard analytical protocols were followed in the analysis of the samples and no problems were encountered or anomalies observed. In addition all laboratory quality control samples were within established control limits, with any exceptions noted below. Each sample was analyzed to achieve the lowest possible reporting limit within the constraints of the method. In some cases, due to interference or analytes present at high concentrations, samples were diluted. For diluted samples, the reporting limits are adjusted relative to the dilution required.

Calculations are performed before rounding to avoid round-off errors in calculated results.

All holding times were met and proper preservation noted for the methods performed on these samples, unless otherwise detailed in the individual sections below.

### **RECEIPT**

The samples were received on 6/29/2018 8:50 AM; the samples arrived in good condition, properly preserved and, where required, on ice. The temperatures of the 8 coolers at receipt time were 0.5° C, 0.6° C, 0.7° C, 0.9° C, 1.0° C, 1.8° C, 2.6° C and 2.7° C.

### **Receipt Exceptions**

One of one HCl preserved VOA vials for the requested 8260B VOCs analysis for TB-062818-01 (280-111519-3) were received with a headspace bubble greater than 6mm in diameter. It can be noted that analytical results may be biased low due to headspace. The laboratory will proceed with the requested analysis unless instructed otherwise. The client was notified on 6/29/2018.

The chain of custody indicates the sample volume received for the requested 6010C/6020A/7470A Dissolved Metals analysis for LL1mw-086-062718-GW (280-111519-7) was field filtered and nitric acid preserved while the laboratory received unpreserved and unfiltered sample volume. The laboratory logged the sample volume for lab filtered and preserved dissolved metals analysis per the volume received. The client was notified on 6/29/2018.

One of three 1L amber glass bottles of unpreserved sample volume for the requested 8270D SVOCs analysis for LL1mw-083-062718-GW (280-111519-14) was received broken. It can be noted that no volume was able to be salvaged by the laboratory and the container was disposed of by the laboratory. Sufficient intact sample volume was received to perform the requested analysis. The client was notified on 6/29/2018.

The sample ID on the container labels of sample volume received for LL1mw-087-062818-GW (280-111519-17) does not match the chain of custody. The container labels list a sample ID of LL1mw-087-062818-GW while the chain of custody lists a sample ID of LL1mw-087-062818. The laboratory logged the sample ID per the container labels which matches the sample name format of all other samples collected. The client was notified on 6/29/2018.

The sample ID on the container labels of sample volume received for SCFmw-004-062818-GW (280-111519-18) does not match the chain of custody. The container labels list a sample ID of SCFmw-004-062818-GW while the chain of custody lists a sample ID of SCFmw-004-062818. The laboratory logged the sample ID per the container labels which matches the sample name format of all other samples collected. The client was notified on 6/29/2018.

The chain of custody indicates the sample volume received for the requested 6010C/6020A/7470A Dissolved Metals analysis for LL12mw-242-D-062718-GW (280-111519-19) was field filtered and nitric acid preserved while the laboratory received unpreserved and unfiltered sample volume. The laboratory logged the sample volume for lab filtered and preserved dissolved metals analysis per the volume received. The client was notified on 6/29/2018.

The chain of custody indicates the sample volume received for the requested 6010C/6020A/7470A Dissolved Metals analysis for LL12mw-242-062718-GW (280-111519-20) was field filtered and nitric acid preserved while the laboratory received unpreserved and unfiltered sample volume. The laboratory logged the sample volume for lab filtered and preserved dissolved metals analysis per the volume received. The client was notified on 6/29/2018.

The requested 8330 Nitroguanidine and 353.2 Nitrocellulose analyses were subcontracted to TestAmerica's Sacramento laboratory.

The sample volumes subcontracted to TestAmerica Sacramento for 8330 Nitroguanidine and 353.2 Nitrocellulose were received at the laboratory outside the required temperature criteria at an elevated temperature of 9.8° C, due to a FedEx shipping delay.

SDG 280-111519-2 was created to report nitrate and nitrite in accordance with method 9056A for samples LL1mw-083-062718-GW (280-111519-14) and RQLmw-011-062818-GW (280-111519-16) as requested by the client on September 26, 2018.

**ANIONS (48 HOURS)**

Samples LL1mw-083-062718-GW (280-111519-14) and RQLmw-011-062818-GW (280-111519-16) were analyzed for anions (48 hours) in accordance with 9056A. The samples were analyzed on 07/21/2018.

The request to report nitrate and nitrite results for samples LL1mw-083-062718-GW (280-111519-14) and RQLmw-011-062818-GW (280-111519-16) was made after the holding times had expired .

No analytical or quality issues were noted, other than those described above or in the Definitions/Glossary page.

# Detection Summary

Client: Cardno GS, Inc  
Project/Site: Ravenna, OH

TestAmerica Job ID: 280-111519-2

**Client Sample ID: LL1mw-083-062718-GW**

**Lab Sample ID: 280-111519-14**

Analyte	Result	Qualifier	LOQ	LOD	DL	Unit	Dil Fac	D	Method	Prep Type
Nitrate as N	310	J H	500	100	42	ug/L	1		9056A	Total/NA

**Client Sample ID: RQLmw-011-062818-GW**

**Lab Sample ID: 280-111519-16**

No Detections.

This Detection Summary does not include radiochemical test results.

TestAmerica Denver

# Client Sample Results

Client: Cardno GS, Inc  
Project/Site: Ravenna, OH

TestAmerica Job ID: 280-111519-2

**Client Sample ID: LL1mw-083-062718-GW**

**Lab Sample ID: 280-111519-14**

Matrix: Water

Date Collected: 06/27/18 13:25  
Date Received: 06/29/18 08:50

## General Chemistry

Analyte	Result	Qualifier	LOQ	LOD	DL	Unit	D	Analyzed	Dil Fac
Nitrate as N	310	J H	500	100	42	ug/L		07/21/18 00:46	1
Nitrite as N	100	U H	500	100	49	ug/L		07/21/18 00:46	1

**Client Sample ID: RQLmw-011-062818-GW**

**Lab Sample ID: 280-111519-16**

Matrix: Water

Date Collected: 06/28/18 11:51  
Date Received: 06/29/18 08:50

## General Chemistry

Analyte	Result	Qualifier	LOQ	LOD	DL	Unit	D	Analyzed	Dil Fac
Nitrate as N	100	U H	500	100	42	ug/L		07/21/18 01:04	1
Nitrite as N	100	U H	500	100	49	ug/L		07/21/18 01:04	1

# Default Detection Limits

Client: Cardno GS, Inc  
Project/Site: Ravenna, OH

TestAmerica Job ID: 280-111519-2

## General Chemistry

Analyte	LOQ	DL	Units	Method
Nitrate as N	500	42	ug/L	9056A
Nitrite as N	500	49	ug/L	9056A

# QC Sample Results

Client: Cardno GS, Inc  
Project/Site: Ravenna, OH

TestAmerica Job ID: 280-111519-2

## Method: 9056A - Anions, Ion Chromatography

**Lab Sample ID:** MB 280-422998/13

**Matrix:** Water

**Analysis Batch:** 422998

Analyte	MB Result	MB Qualifier	LOQ	LOD	DL	Unit	D	Analyzed	Dil Fac
Nitrate as N	100	U	500	100	42	ug/L		07/20/18 13:47	1
Nitrite as N	100	U	500	100	49	ug/L		07/20/18 13:47	1

**Lab Sample ID:** LCS 280-422998/11

**Matrix:** Water

**Analysis Batch:** 422998

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec. Limits		
Nitrate as N	5000	4980		ug/L		100	88 - 111		
Nitrite as N	5000	5120		ug/L		102	87 - 111		

**Lab Sample ID:** LCSD 280-422998/12

**Matrix:** Water

**Analysis Batch:** 422998

Analyte	Spike Added	LCSD Result	LCSD Qualifier	Unit	D	%Rec	%Rec. Limits	RPD	Limit
Nitrate as N	5000	4980		ug/L		100	88 - 111	0	10
Nitrite as N	5000	5120		ug/L		102	87 - 111	0	10

**Lab Sample ID:** MRL 280-422998/16

**Matrix:** Water

**Analysis Batch:** 422998

Analyte	Spike Added	MRL Result	MRL Qualifier	Unit	D	%Rec	%Rec. Limits		
Nitrate as N	0.200	0.223	J	mg/L		111	50 - 150		
Nitrite as N	0.200	0.223	J	mg/L		111	50 - 150		

**Lab Sample ID:** 280-111519-16 MS

**Matrix:** Water

**Analysis Batch:** 422998

Analyte	Sample Result	Sample Qualifier	Spike Added	MS Result	MS Qualifier	Unit	D	%Rec	%Rec. Limits		
Nitrate as N	100	U H	5000	4870	H	ug/L		97	88 - 111		
Nitrite as N	100	U H	5000	4650	H	ug/L		93	87 - 111		

**Lab Sample ID:** 280-111519-16 MSD

**Matrix:** Water

**Analysis Batch:** 422998

Analyte	Sample Result	Sample Qualifier	Spike Added	MSD Result	MSD Qualifier	Unit	D	%Rec	%Rec. Limits	RPD	Limit
Nitrate as N	100	U H	5000	4920	H	ug/L		98	88 - 111	1	10
Nitrite as N	100	U H	5000	4710	H	ug/L		94	87 - 111	1	10

**Lab Sample ID:** 280-111519-16 DU

**Matrix:** Water

**Analysis Batch:** 422998

Analyte	Sample Result	Sample Qualifier		DU Result	DU Qualifier	Unit	D		RPD	Limit
Nitrate as N	100	U H		100	U	ug/L			NC	10

TestAmerica Denver

# QC Sample Results

Client: Cardno GS, Inc  
Project/Site: Ravenna, OH

TestAmerica Job ID: 280-111519-2

## Method: 9056A - Anions, Ion Chromatography (Continued)

Lab Sample ID: 280-111519-16 DU

Client Sample ID: RQLmw-011-062818-GW

Matrix: Water

Prep Type: Total/NA

Analysis Batch: 422998

Analyte	Sample	Sample	DU	DU	Unit	D	RPD	RPD
	Result	Qualifier	Result	Qualifier				
Nitrite as N	100	U H	100	U	ug/L		NC	10

# QC Association Summary

Client: Cardno GS, Inc  
Project/Site: Ravenna, OH

TestAmerica Job ID: 280-111519-2

## General Chemistry

### Analysis Batch: 422998

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
280-111519-14	LL1mw-083-062718-GW	Total/NA	Water	9056A	
280-111519-16	RQLmw-011-062818-GW	Total/NA	Water	9056A	
MB 280-422998/13	Method Blank	Total/NA	Water	9056A	
LCS 280-422998/11	Lab Control Sample	Total/NA	Water	9056A	
LCSD 280-422998/12	Lab Control Sample Dup	Total/NA	Water	9056A	
MRL 280-422998/16	Lab Control Sample	Total/NA	Water	9056A	
280-111519-16 MS	RQLmw-011-062818-GW	Total/NA	Water	9056A	
280-111519-16 MSD	RQLmw-011-062818-GW	Total/NA	Water	9056A	
280-111519-16 DU	RQLmw-011-062818-GW	Total/NA	Water	9056A	

# Lab Chronicle

Client: Cardno GS, Inc  
Project/Site: Ravenna, OH

TestAmerica Job ID: 280-111519-2

**Client Sample ID: LL1mw-083-062718-GW**

**Lab Sample ID: 280-111519-14**

**Matrix: Water**

Date Collected: 06/27/18 13:25  
Date Received: 06/29/18 08:50

Prep Type	Batch Type	Batch Method	Run	Dil Factor	Initial Amount	Final Amount	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	9056A		1	5 mL	5 mL	422998	07/21/18 00:46	CCJ	TAL DEN

**Client Sample ID: RQLmw-011-062818-GW**

**Lab Sample ID: 280-111519-16**

**Matrix: Water**

Date Collected: 06/28/18 11:51  
Date Received: 06/29/18 08:50

Prep Type	Batch Type	Batch Method	Run	Dil Factor	Initial Amount	Final Amount	Batch Number	Prepared or Analyzed	Analyst	Lab
Total/NA	Analysis	9056A		1	5 mL	5 mL	422998	07/21/18 01:04	CCJ	TAL DEN

**Laboratory References:**

TAL DEN = TestAmerica Denver, 4955 Yarrow Street, Arvada, CO 80002, TEL (303)736-0100

# Accreditation/Certification Summary

Client: Cardno GS, Inc  
Project/Site: Ravenna, OH

TestAmerica Job ID: 280-111519-2

## Laboratory: TestAmerica Denver

The accreditations/certifications listed below are applicable to this report.

Authority	Program	EPA Region	Identification Number	Expiration Date
A2LA	DoD ELAP		2907.01	10-31-19

# Method Summary

Client: Cardno GS, Inc  
Project/Site: Ravenna, OH

TestAmerica Job ID: 280-111519-2

Method	Method Description	Protocol	Laboratory
9056A	Anions, Ion Chromatography	SW846	TAL DEN

**Protocol References:**

SW846 = "Test Methods For Evaluating Solid Waste, Physical/Chemical Methods", Third Edition, November 1986 And Its Updates.

**Laboratory References:**

TAL DEN = TestAmerica Denver, 4955 Yarrow Street, Arvada, CO 80002, TEL (303)736-0100

# Sample Summary

Client: Cardno GS, Inc  
Project/Site: Ravenna, OH

TestAmerica Job ID: 280-111519-2

Lab Sample ID	Client Sample ID	Matrix	Collected	Received
280-111519-14	LL1mw-083-062718-GW	Water	06/27/18 13:25	06/29/18 08:50
280-111519-16	RQLmw-011-062818-GW	Water	06/28/18 11:51	06/29/18 08:50

## GENERAL CHEMISTRY MANUAL INTEGRATION SUMMARY

Lab Name: TestAmerica DenverJob No.: 280-111519-2

SDG No.: \_\_\_\_\_

Instrument ID: WC\_IonChrom7Analysis Batch Number: 422998Lab Sample ID: CCV 280-422998/14

Client Sample ID: \_\_\_\_\_

Date Analyzed: 07/20/18 14:05Lab File ID: 14.0000.dGC Column: Ion PAC AS 17 ID: \_\_\_\_\_

COMPOUND NAME	RETENTION TIME	MANUAL INTEGRATION		
		REASON	ANALYST	DATE
Sulfate	8.93	Wrong Peak	jewellc	07/20/18 14:55

## REAGENT TRACEABILITY SUMMARY

Lab Name: TestAmerica Denver

Job No.: 280-111519-2

SDG No.:

Reagent ID	Exp Date	Prep Date	Dilutant Used	Reagent Final Volume	Parent Reagent		Analyte	Concentration
					Reagent ID	Volume Added		
<b>IC Cal low_00373</b>	06/11/18	06/04/18	Di Water, Lot NA	100 mL	IC N02 CAL_00042	5 mL	Nitrite as N	50 mg/L
					IC N03 cal_00018	5 mL	Nitrate as N	50 mg/L
					IC P04 cal_00018	5 mL	Orthophosphate as P	50 mg/L
.IC N02 CAL_00042	08/31/18		RICCA, Lot 1802e42		(Purchased Reagent)		Nitrite as N	1000 ppm
.IC N03 cal_00018	11/30/18		Ricca, Lot 2705D50		(Purchased Reagent)		Nitrate as N	1000 mg/L
.IC P04 cal_00018	11/30/19		RICCA, Lot 4711L59		(Purchased Reagent)		Orthophosphate as P	1000 mg/L
<b>IC Cal low_00385</b>	07/24/18	07/17/18	Di Water, Lot NA	100 mL	IC N02 CAL_00042	5 mL	Nitrite as N	50 mg/L
					IC N03 cal_00018	5 mL	Nitrate as N	50 mg/L
					(Purchased Reagent)		Nitrite as N	1000 ppm
.IC N02 CAL_00042	08/31/18		RICCA, Lot 1802e42		(Purchased Reagent)		Nitrite as N	1000 ppm
.IC N03 cal_00018	11/30/18		Ricca, Lot 2705D50		(Purchased Reagent)		Nitrate as N	1000 mg/L
<b>IC ICV 5_00201</b>	06/06/18	05/30/18	Di Water, Lot na	10 mL	IC N02 ICV 00015	0.5 mL	Nitrite as N	50 mg/L
					IC N03 ICV 00012	0.5 mL	Nitrate as N	50 mg/L
					(Purchased Reagent)		Nitrite as N	1000 mg/L
.IC N02 ICV 00015	06/30/18		ERA, Lot 320616		(Purchased Reagent)		Nitrite as N	1000 mg/L
.IC N03 ICV 00012	12/31/18		ERA, Lot 140616		(Purchased Reagent)		Nitrate as N	1000 mg/L
<b>IC LCS_01288</b>	07/21/18	07/20/18	Di Water, Lot 27	200 mL	IC Cal low_00386	20 mL	Nitrite as N	5 mg/L
							Nitrate as N	5 mg/L
					IC N02 CAL_00042	5 mL	Nitrite as N	50 mg/L
.IC Cal low_00386	07/25/18	07/18/18	Di Water, Lot NA	100 mL	IC N03 cal_00018	5 mL	Nitrate as N	50 mg/L
..IC N02 CAL_00042	08/31/18		RICCA, Lot 1802e42		(Purchased Reagent)		Nitrite as N	1000 ppm
..IC N03 cal_00018	11/30/18		Ricca, Lot 2705D50		(Purchased Reagent)		Nitrate as N	1000 mg/L
<b>ICMS/MSD WEEK_00543</b>	07/26/18	07/19/18	Di Water, Lot NA	10 mL	IC SPK 6 ANIO_00019	5 mL	Nitrate as N	500.003 mg/L
					IC SPK N02SOL 00013	5 mL	Nitrite as N	499.973 mg/L
					IC MS/MSD N03_00004	6.068 g	Nitrate as N	1000.01 mg/L
..IC SPK 6 ANIO_00019	08/23/18	08/23/17	Di Water, Lot NA	1000 mL	(Purchased Reagent)		Nitrate as N	0.1648 g/g
..IC MS/MSD N03_00004	10/02/18		FISHER, Lot 035600		(Purchased Reagent)		Nitrate as N	0.1648 g/g
.IC SPK N02SOL 00013	06/05/19	06/05/18	Di Water, Lot na	500 mL	IC MS/MSD N02_00002	2.4628 g	Nitrite as N	999.946 mg/L
..IC MS/MSD N02_00002	01/12/27		fisher, Lot 164254		(Purchased Reagent)		Nitrite as N	0.20301 g/g

Reagent

---

**IC MS/MSD N02\_00002**



1 Reagent Lane  
Fair Lawn, NJ 07410  
201.796.7100 tel  
201.796.1329 fax


## Certificate of Analysis

Fisher Scientific's Quality System has been found to conform to Quality Management System Standard ISO9001:2008 standard by SAI Global Certificate Number CERT - 0090918

This is to certify that units of the lot number below were tested and found to comply with the specifications of the grade listed. Certain data have been supplied by third parties. Fisher Scientific expressly disclaims all warranties, expressed or implied, including the implied warranties of merchantability and fitness for a particular purpose. Certain products (USP/FCC/NF/EP/BP/JP grades) are sold for use in food, drug, or medical device manufacturing. Fisher does not claim regulatory coverage under 21 CFR nor maintain DMP's with the FDA. The following are the actual analytical results obtained:

Catalog Number	S347	Quality Test / Release Date	9/9/2016
Lot Number	164254		
Description	SODIUM NITRITE, A.C.S.		
Country of Origin	India	* Suggested Retest Date	Sep-2021
Chemical Origin	Inorganic-non animal		
BSE/TSE Comment	No animal products are used as starting raw material ingredients, or used in processing, including lubricants, processing aids, or any other material that might migrate to the finished product.		

Result name	Units	Specifications	Test Value
APPEARANCE		REPORT	YELLOW-WHITE CRYSTALS
ASSAY	%	>= 97	98.6
CALCIUM	%	<= 0.01	<0.010
CHLORIDE	%	<= 0.005	<0.005
HEAVY METALS (as Pb)	%	<= 0.001	<0.0010
IDENTIFICATION	PASS/FAIL	= PASS TEST	PASS TEST
INSOLUBLE MATTER	%	<= 0.01	<0.010
IRON (Fe)	%	<= 0.001	<0.0010
POTASSIUM (K)	%	<= 0.005	<0.0010
SULFATE (SO4)	%	<= 0.01	<0.010



CERTIFIED BY  
Teresa Bailey-Wyche  
Quality Assurance Specialist - Certificate of Analysis Fair Lawn

Note: The data listed is valid for all package sizes of this lot of this product, expressed as a extension of this catalog number listed above. If there are any questions with this certificate, please call Chemical Services at (800) 227-6701.  
\*Based on suggested storage condition.

Reagent

---

**IC N02 CAL\_00042**



# Certificate of Analysis

## Nitrite Nitrogen Standard, 1000 ppm N (3285 ppm NO<sub>2</sub>)

Lot Number: 1802E42

Product Number: R5444900

Manufacture Date: FEB 14, 2018

Expiration Date: AUG 2018

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Potassium Nitrite	7758-09-0	ACS

Test	Specification	Result	NIST SRM#
Appearance	Colorless liquid	Passed	
Assay (vs. Potassium Permanganate)	995-1005 ppm N	1001 ppm N	40

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
R5444900-120C	120 mL amber glass	6 months
R5444900-500C	500 mL amber glass	6 months

**Recommended Storage:** 2°C - 8°C (36°F - 46°F)

Israel Alamudun (02/14/2018)

Quality Control Supervisor

This Certificate of Analysis is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.

Reagent

---

**IC N03 cal\_00018**



# Certificate of Analysis

## Nitrate Nitrogen Standard, 1000 ppm N (4427 ppm NO<sub>3</sub>)

Lot Number: 2705D50

Product Number: 5459

Manufacture Date: MAY 24, 2017

Expiration Date: NOV 2018

The certified value reported is the prepared value based upon the method of preparation of the material. The uncertainty in the prepared value is based upon the volumetric method of preparation.

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Potassium Nitrate	7757-79-1	High Purity
Chloroform	67-66-3	

Test	Specification	Result
Appearance	Colorless liquid	Passed
Nitrogen (N)	995-1005 ppm	1000 ppm

Specification	Reference
Nitrate Solution, Stock (1.0 mL = 1.0 mg NO <sub>3</sub> -N)	ASTM (D 3867 A)
Nitrate Solution, Stock (1.0 mL = 1.0 mg NO <sub>3</sub> -N)	ASTM (D 3867 B)
Stock Nitrate Solution: 1 mL = 1.0 mg NO <sub>3</sub> -N	EPA (353.2)
Stock Nitrate Solution: 1.0 mL = 1.00 mg NO <sub>3</sub> -N	EPA (353.3)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
5459-16	500 mL natural poly	18 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

Andy Baumgartner (05/24/2017)

Quality Control Supervisor

This Certificate of Analysis is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

Reagent

---

**IC NO2 ICV\_00015**



A Waters Company

## Certificate of Analysis

PRODUCT:	1000 mg/L Nitrite as N (NO <sub>2</sub> -N)
CATALOG NUMBER:	053 -125 mL; 990 - 500 mL
LOT NUMBER:	320616
ISSUE DATE:	July 7, 2016
REVISION DATE:	Original
STARTING MATERIAL:	Sodium Nitrite (NaNO <sub>2</sub> )
CERTIFIED CONCENTRATION <sup>1</sup> :	1000 mg/L
UNCERTAINTY <sup>2</sup> :	0.9%
MATRIX:	18 megohm deionized water
DENSITY:	1.0001 ± 0.0016 g/mL at 20.0°C and 761 mm Hg
TRACEABILITY <sup>3</sup> :	NA
NIST/SRM:	SRM not available
VERIFICATION METHOD:	Ion Chromatography
STORAGE:	Store at 20-25°C

1. The Certified Concentration is the actual made-to concentration confirmed by ERA analytical verification.
2. The stated Uncertainty is the total propagated uncertainty at the 95% confidence interval. The uncertainty is based on the preparation and internal analytical verification of the product by ERA, multiplied by a coverage factor which is equal to the student t factor at a 95% confidence interval at n-1 degrees of freedom. The uncertainty applies to the product as supplied and does not take into account any required or optional dilutions and/or preparations the laboratory may perform while using this product.
3. Traceability Recovery = ((% Recovery certified standard)/(% Recovery NIST SRM))\*100.

The traceability data shown were compiled by analyzing the ERA standards or their associated stock solutions against the applicable NIST SRMs.

---

This standard **expires 6/2018**. The certified values are monitored and purchasers will be notified of any significant changes resulting in recertification or withdrawal of this certified reference material during the period of validity of this certificate.

This product is intended to be used as either a calibration standard or a quality control check of the entire analytical process for the analytes/matrix included in the standard.

If you have any questions or need technical assistance, please call ERA technical assistance at 1-800-372-0122 or email to info@eraqc.com

Certifying Officer: Brian Miller

ISO/IEC GUIDE 34:2009

ISO/IEC 17025:2005



REFERENCE MATERIAL PRODUCER  
CERTIFICATE NO. 1539.03



CHEMICAL TESTING LABORATORY  
CERTIFICATE NO. 1539.02

Reagent

---

**IC NO3 ICV\_00012**

# Certificate of Analysis

**PRODUCT:** 1000 mg/L Nitrate as N (NO<sub>3</sub>-N)  
**CATALOG NUMBER:** 052 -125 mL; 991 - 500 mL  
**LOT NUMBER:** 140616  
**ISSUE DATE:** June 30, 2016  
**REVISION DATE:** June 26, 2017  
  
**STARTING MATERIAL:** Potassium Nitrate (KNO<sub>3</sub>)  
**CERTIFIED CONCENTRATION<sup>1</sup>:** 1000 mg/L  
**UNCERTAINTY<sup>2</sup>:** 0.6%  
**MATRIX:** 18 megohm deionized water  
**DENSITY:** 1.0020 ± 0.0008 g/mL at 21.5°C and 762 mm Hg  
  
**TRACEABILITY<sup>3</sup>:** 102%  
**NIST/SRM:** 3185 Nitrate  
**VERIFICATION METHOD:** Ion Chromatography  
**STORAGE:** Store at 20-25°C

1. The **Certified Concentration** is the actual made-to concentration confirmed by ERA analytical verification.
2. The stated **Uncertainty** is the total propagated uncertainty at the 95% confidence interval. The uncertainty is based on the preparation and internal analytical verification of the product by ERA, multiplied by a coverage factor which is equal to the student t factor at a 95% confidence interval at n-1 degrees of freedom. The uncertainty applies to the product as supplied and does not take into account any required or optional dilutions and/or preparations the laboratory may perform while using this product.
3. Traceability Recovery = ((% Recovery certified standard)/(% Recovery NIST SRM))\*100.

The traceability data shown were compiled by analyzing the ERA standards or their associated stock solutions against the applicable NIST SRMs.

This standard **expires 12/2018**. The certified values are monitored and purchasers will be notified of any significant changes resulting in recertification or withdrawal of this certified reference material during the period of validity of this certificate.

This product is intended to be used as either a calibration standard or a quality control check of the entire analytical process for the analytes/matrix included in the standard.

If you have any questions or need technical assistance, please call ERA technical assistance at 1-800-372-0122 or email to info@eraqc.com

Certifying Officer: Brian Miller

ISO/IEC GUIDE 34:2009



REFERENCE MATERIAL PRODUCER  
CERTIFICATE NO. 1539.03

ISO/IEC 17025:2005



CHEMICAL TESTING LABORATORY  
CERTIFICATE NO. 1539.02

Reagent

---

**IC P04 cal\_00018**



# Certificate of Analysis

## Phosphorus AA Standard, 1000 ppm P in H<sub>2</sub>O

Lot Number: 4711L59

Product Number: AP1KW

Manufacture Date: NOV 30, 2017

Expiration Date: NOV 2019

This is a single element solution that was prepared volumetrically to contain the certified value reported. The uncertainty associated with the certified value is the sum of the estimated errors due to the purity of the raw material, the volumetric preparation of the solution, and transpiration of the solution through the container wall.

The final solution concentration is confirmed by AA, ICP, or ICP-MS, and is traceable to NIST Standard Reference Material 3139.


Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Ammonium Dihydrogen Phosphate	7722-76-1	High Purity

Test	Specification	Result	NIST SRM#
Appearance	Colorless liquid	Passed	
Phosphorus (P)	995-1005 ppm	1000 ppm	3139

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
AP1KW-500	500 mL natural poly	24 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

  
Jim Gibbs (11/30/2017)

Quality Control Supervisor

This Certificate of Analysis is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

# **GENERAL CHEMISTRY**

COVER PAGE  
GENERAL CHEMISTRY

Lab Name: TestAmerica Denver Job Number: 280-111519-2

SDG No.: \_\_\_\_\_

Project: Ravenna, OH

Client Sample ID  
LL1mw-083-062718-GW  
RQLmw-011-062818-GW

Lab Sample ID  
280-111519-14  
280-111519-16

Comments:

1B-IN  
INORGANIC ANALYSIS DATA SHEET  
GENERAL CHEMISTRY

Client Sample ID: LL1mw-083-062718-GW

Lab Sample ID: 280-111519-14

Lab Name: TestAmerica Denver

Job No.: 280-111519-2

SDG ID.:

Matrix: Water

Date Sampled: 06/27/2018 13:25

Reporting Basis: WET

Date Received: 06/29/2018 08:50

Analyte	Result	LOQ	LOD	DL	Units	C	Q	DIL	Method
Nitrate as N	310	500	100	42	ug/L	J	H	1	9056A
Nitrite as N	100	500	100	49	ug/L	U	H	1	9056A

1B-IN  
INORGANIC ANALYSIS DATA SHEET  
GENERAL CHEMISTRY

Client Sample ID: RQLmw-011-062818-GW

Lab Sample ID: 280-111519-16

Lab Name: TestAmerica Denver

Job No.: 280-111519-2

SDG ID.:

Matrix: Water

Date Sampled: 06/28/2018 11:51

Reporting Basis: WET

Date Received: 06/29/2018 08:50

Analyte	Result	LOQ	LOD	DL	Units	C	Q	DIL	Method
Nitrate as N	100	500	100	42	ug/L	U	H	1	9056A
Nitrite as N	100	500	100	49	ug/L	U	H	1	9056A

2-IN  
CALIBRATION QUALITY CONTROL  
GENERAL CHEMISTRY

Lab Name: TestAmerica Denver Job No.: 280-111519-2

SDG No.: \_\_\_\_\_

Analyst: CCJ Batch Start Date: 06/05/2018

Reporting Units: mg/L Analytical Batch No.: 417404

Sample Number	QC Type	Time	Analyte	Result	Spike Amount	(%) Recovery	Limits	Qual	Reagent
8	ICV	12:20	Nitrate as N	4.10	4.00	103	90-110	U	IC ICV 5_00201
			Nitrite as N	4.15	4.00	104	90-110		IC ICV 5_00201
9	ICB	12:38	Nitrate as N	0.10				U	
			Nitrite as N	0.10					

Note! Calculations are performed before rounding to avoid round-off errors in calculated results.

FORM II-IN

2-IN  
CALIBRATION QUALITY CONTROL  
GENERAL CHEMISTRY

Lab Name: TestAmerica Denver Job No.: 280-111519-2  
SDG No.: \_\_\_\_\_  
Analyst: CCJ Batch Start Date: 07/20/2018  
Reporting Units: mg/L Analytical Batch No.: 422998

Sample Number	QC Type	Time	Analyte	Result	Spike Amount	(%) Recovery	Limits	Qual	Reagent
1	CCV	10:14	Nitrate as N	5.01	5.00	100	90-110	IC	LCS_01288
			Nitrite as N	5.12	5.00	102	90-110	IC	LCS_01288
2	CCB	10:32	Nitrate as N	0.10				U	
			Nitrite as N	0.10				U	
14	CCV	14:05	Nitrate as N	4.97	5.00	99	90-110	IC	LCS_01288
			Nitrite as N	5.12	5.00	102	90-110	IC	LCS_01288
15	CCB	14:23	Nitrate as N	0.10				U	
			Nitrite as N	0.10				U	
38	CCV	00:10	Nitrate as N	5.12	5.00	102	90-110	IC	LCS_01288
			Nitrite as N	5.14	5.00	103	90-110	IC	LCS_01288
39	CCB	00:28	Nitrate as N	0.10				U	
			Nitrite as N	0.10				U	
50	CCV	03:44	Nitrate as N	5.08	5.00	102	90-110	IC	LCS_01288
			Nitrite as N	5.15	5.00	103	90-110	IC	LCS_01288
51	CCB	04:02	Nitrate as N	0.10				U	
			Nitrite as N	0.10				U	

Note! Calculations are performed before rounding to avoid round-off errors in calculated results.

FORM II-IN

3-IN  
METHOD BLANK  
GENERAL CHEMISTRY

Lab Name: TestAmerica Denver Job No.: 280-111519-2

SDG No.: \_\_\_\_\_

Method	Lab Sample ID	Analyte	Result	Qual	Units	LOQ	Dil
Batch ID: 422998 Date: 07/20/2018 13:47							
9056A	MB 280-422998/13	Nitrate as N	100	U	ug/L	500	1
9056A	MB 280-422998/13	Nitrite as N	100	U	ug/L	500	1

5-IN  
MATRIX SPIKE SAMPLE RECOVERY  
GENERAL CHEMISTRY

Lab Name: TestAmerica Denver Job No.: 280-111519-2

SDG No.: \_\_\_\_\_

Matrix: Water

Method	Lab Sample ID	Analyte	Result	C	Unit	Spike Amount	Pct. Rec.	Limits	RPD	RPD Limit	Q
Batch ID: 422998 Date: 07/21/2018 01:39											
9056A	280-111519-16	Nitrate as N	100	U	ug/L						H
9056A	280-111519-16	Nitrate as N MS	4870		ug/L	5000	97	88-111			H
9056A	280-111519-16	Nitrite as N	100	U	ug/L						H
9056A	280-111519-16	Nitrite as N MS	4650		ug/L	5000	93	87-111			H

Calculations are performed before rounding to avoid round-off errors in calculated results.

5-IN  
MATRIX SPIKE DUPLICATE SAMPLE RECOVERY  
GENERAL CHEMISTRY

Lab Name: TestAmerica Denver Job No.: 280-111519-2

SDG No.: \_\_\_\_\_

Matrix: Water

Method	Lab Sample ID	Analyte	Result	C	Unit	Spike Amount	Pct. Rec.	Limits	RPD	RPD Limit	Q
Batch ID: 422998 Date: 07/21/2018 01:57											
9056A	280-111519-16	Nitrate as N	4920		ug/L	5000	98	88-111	1	10	H
9056A	280-111519-16	Nitrite as N	4710		ug/L	5000	94	87-111	1	10	H

Calculations are performed before rounding to avoid round-off errors in calculated results.

6-IN  
DUPLICATE  
GENERAL CHEMISTRY

Lab Name: TestAmerica Denver Job No.: 280-111519-2

SDG No.: \_\_\_\_\_

Matrix: Water

Method	Client Sample ID	Lab Sample ID	Analyte	Result	Unit	RPD	RPD Limit	Qual
Batch ID: 422998 Date: 07/21/2018 01:21								
9056A	RQLmw-011-062818-G	280-111519-16	Nitrate as N W	100	ug/L			U
9056A	RQLmw-011-062818-G	280-111519-16	Nitrate as N W	100	ug/L	NC	10	U
9056A	RQLmw-011-062818-G	280-111519-16	Nitrite as N W	100	ug/L			U
9056A	RQLmw-011-062818-G	280-111519-16	Nitrite as N W	100	ug/L	NC	10	U

Calculations are performed before rounding to avoid round-off errors in calculated results.

FORM VI-IN

7A-IN  
LAB CONTROL SAMPLE  
GENERAL CHEMISTRY

Lab Name: TestAmerica Denver

Job No.: 280-111519-2

SDG No.: \_\_\_\_\_

Matrix: Water

Method	Lab Sample ID	Analyte	Result	C	Unit	Spike Amount	Pct. Rec.	Limits	RPD	RPD Limit	Q
Batch ID: 422998 Date: 07/20/2018 13:12											
LCS Source: IC LCS_01288											
9056A	LCS 280-422998/11	Nitrate as N	4980		ug/L	5000	100	88-111	0	10	
9056A	LCS 280-422998/11	Nitrite as N	5120		ug/L	5000	102	87-111	0	10	

Calculations are performed before rounding to avoid round-off errors in calculated results.

FORM VIIA-IN

7A-IN  
LAB CONTROL SAMPLE DUPLICATE  
GENERAL CHEMISTRY

Lab Name: TestAmerica Denver Job No.: 280-111519-2

SDG No.: \_\_\_\_\_

Matrix: Water

Method	Lab Sample ID	Analyte	Result	C	Unit	Spike Amount	Pct. Rec.	Limits	RPD	RPD Limit	Q
Batch ID: 422998 Date: 07/20/2018 13:30 LCSD Source: IC LCS_01288											
9056A	LCSD 280-422998/12	Nitrate as N	4980		ug/L	5000	100	88-111	0	10	
9056A	LCSD 280-422998/12	Nitrite as N	5120		ug/L	5000	102	87-111	0	10	

Calculations are performed before rounding to avoid round-off errors in calculated results.

FORM VIIA-IN

7A-IN  
METHOD REPORTING LIMIT CHECK  
GENERAL CHEMISTRY

Lab Name: TestAmerica Denver

Job No.: 280-111519-2

SDG No.: \_\_\_\_\_

Matrix: Water

Method	Lab Sample ID	Analyte	Result	C	Unit	Spike Amount	Pct. Rec.	Limits	RPD	RPD Limit	Q
Batch ID: 422998 Date: 07/20/2018 17:38											
LCS Source: IC Cal low_00385											
9056A	MRL 280-422998/16	Nitrate as N	0.223	J	mg/L	0.200	111	50-150			
9056A	MRL 280-422998/16	Nitrite as N	0.223	J	mg/L	0.200	111	50-150			

Calculations are performed before rounding to avoid round-off errors in calculated results.

FORM VIIA-IN

9-IN  
DETECTION LIMITS  
GENERAL CHEMISTRY

Lab Name: TestAmerica Denver

Job Number: 280-111519-2

SDG Number: \_\_\_\_\_

Matrix: Water

Instrument ID: WC\_IonChrom7

Method: 9056A

DL Date: 02/16/2014 00:00

Analyte	Wavelength/ Mass	LOQ (mg/L)	DL (mg/L)
Nitrate as N		0.5	0.042
Nitrite as N		0.5	0.049

9-IN  
CALIBRATION BLANK DETECTION LIMITS  
GENERAL CHEMISTRY

Lab Name: TestAmerica Denver

Job Number: 280-111519-2

SDG Number: \_\_\_\_\_

Matrix: Water

Instrument ID: WC\_IonChrom7

Method: 9056A

XMDL Date: 02/16/2014 00:00

Analyte	Wavelength/ Mass	XRL (mg/L)	XMDL (mg/L)
Nitrate as N		0.5	0.042
Nitrite as N		0.5	0.049

13-IN  
ANALYSIS RUN LOG  
GENERAL CHEMISTRY

Lab Name: TestAmerica Denver

Job No.: 280-111519-2

SDG No.:

Instrument ID: WC\_IonChrom7

Analysis Method: 9056A

Start Date: 06/05/2018 10:15

End Date: 06/06/2018 08:48

13-IN  
ANALYSIS RUN LOG  
GENERAL CHEMISTRY

Lab Name: TestAmerica Denver Job No.: 280-111519-2

SDG No.:

Instrument ID: WC\_IonChrom7 Analysis Method: 9056A

Start Date: 06/05/2018 10:15 End Date: 06/06/2018 08:48

## Prep Types:

13-IN  
ANALYSIS RUN LOG  
GENERAL CHEMISTRY

Lab Name: TestAmerica Denver

Job No.: 280-111519-2

SDG No.:

Instrument ID: WC\_IonChrom7

Analysis Method: 9056A

Start Date: 07/20/2018 10:14

End Date: 07/21/2018 04:02

13-IN  
ANALYSIS RUN LOG  
GENERAL CHEMISTRY

Lab Name: TestAmerica Denver Job No.: 280-111519-2

SDG No.:

Instrument ID: WC\_IonChrom7 Analysis Method: 9056A

Start Date: 07/20/2018 10:14 End Date: 07/21/2018 04:02

## Prep Types:

$$\overline{T} = \text{Total/NA}$$

## GENERAL CHEMISTRY BATCH WORKSHEET

Lab Name: TestAmerica Denver

Job No.: 280-111519-2

SDG No.:

Batch Number: 417404

Batch Start Date: 06/05/18 10:15

Batch Analyst: Jewell, Connie C

Batch Method: 9056A

Batch End Date:

Lab Sample ID	Client Sample ID	Method Chain	Basis	InitialAmount	FinalAmount	IC CAL cl/so4 00202	IC Cal low 00373	IC CL ICV 00014	IC ICV 5 00201
STD 280-417404/2 IC		9056A		5 mL	5 mL	0.02 mL	0.02 mL		
STD 280-417404/3 IC		9056A		5 mL	5 mL	0.05 mL	0.05 mL		
STD 280-417404/4 IC		9056A		5 mL	5 mL	0.1 mL	0.1 mL		
STD 280-417404/5 IC		9056A		5 mL	5 mL	1.2 mL	0.4 mL		
STD 280-417404/6 IC		9056A		5 mL	5 mL	2.4 mL	0.8 mL		
STD 280-417404/7 IC		9056A		5 mL	5 mL	4 mL	1 mL		
ICV 280-417404/8		9056A		5 mL	5 mL			0.4 mL	0.4 mL
ICB 280-417404/9		9056A		5 mL	5 mL				

Lab Sample ID	Client Sample ID	Method Chain	Basis	IC SO4 ICV 00017					
STD 280-417404/2 IC		9056A							
STD 280-417404/3 IC		9056A							
STD 280-417404/4 IC		9056A							
STD 280-417404/5 IC		9056A							
STD 280-417404/6 IC		9056A							
STD 280-417404/7 IC		9056A							
ICV 280-417404/8		9056A		0.4 mL					
ICB 280-417404/9		9056A							

The pound sign (#) in the amount added field denotes that the reagent was used undiluted. All calculations are performed using the stated concentration for this reagent.

## GENERAL CHEMISTRY BATCH WORKSHEET

Lab Name: TestAmerica Denver

Job No.: 280-111519-2

SDG No.:

Batch Number: 417404

Batch Start Date: 06/05/18 10:15

Batch Analyst: Jewell, Connie C

Batch Method: 9056A

Batch End Date:

Batch Notes	
Pipette/Syringe/Dispenser ID	wc5000ccj, wc1000cj, wc200cj
Regeneration Solution ID	171210701013
Sufficient Volume for Batch QC	yes

Basis	Basis Description

The pound sign (#) in the amount added field denotes that the reagent was used undiluted. All calculations are performed using the stated concentration for this reagent.

9056A

Page 2 of 2

## GENERAL CHEMISTRY BATCH WORKSHEET

Lab Name: TestAmerica Denver

Job No.: 280-111519-2

SDG No.:

Batch Number: 422998

Batch Start Date: 07/20/18 10:14

Batch Analyst: Jewell, Connie C

Batch Method: 9056A

Batch End Date:

Lab Sample ID	Client Sample ID	Method Chain	Basis	InitialAmount	FinalAmount	IC CAL cl/s04 00208	IC Cal low 00385	IC LCS 01288	ICMS/MSD WEEK 00543
CCV 280-422998/1		9056A		5 mL	5 mL			5 mL	
CCB 280-422998/2		9056A		5 mL	5 mL				
MRL 280-422998/10		9056A		5 mL	5 mL	0.05 mL	0.02 mL		
LCS 280-422998/11		9056A		5 mL	5 mL			5 mL	
LCSD 280-422998/12		9056A		5 mL	5 mL			5 mL	
MB 280-422998/13		9056A		5 mL	5 mL				
CCV 280-422998/14		9056A		5 mL	5 mL			5 mL	
CCB 280-422998/15		9056A		5 mL	5 mL				
MRL 280-422998/16		9056A		5 mL	5 mL	0.05 mL	0.02 mL		
CCV 280-422998/38		9056A		5 mL	5 mL			5 mL	
CCB 280-422998/39		9056A		5 mL	5 mL				
280-111519-H-14	LL1mw-083-062718 -GW	9056A	T	5 mL	5 mL				
280-111519-D-16	RQLmw-011-062818 -GW	9056A	T	5 mL	5 mL				
280-111519-D-16 DU	RQLmw-011-062818 -GW	9056A	T	5 mL	5 mL				
280-111519-D-16 MS	RQLmw-011-062818 -GW	9056A	T	5 mL	5 mL				0.05 mL
280-111519-D-16 MSD	RQLmw-011-062818 -GW	9056A	T	5 mL	5 mL				0.05 mL
CCV 280-422998/50		9056A		5 mL	5 mL			5 mL	
CCB 280-422998/51		9056A		5 mL	5 mL				

The pound sign (#) in the amount added field denotes that the reagent was used undiluted. All calculations are performed using the stated concentration for this reagent.

## GENERAL CHEMISTRY BATCH WORKSHEET

Lab Name: TestAmerica Denver

Job No.: 280-111519-2

SDG No.:

Batch Number: 422998

Batch Start Date: 07/20/18 10:14

Batch Analyst: Jewell, Connie C

Batch Method: 9056A

Batch End Date:

Batch Notes	
Pipette/Syringe/Dispenser ID	wc5000ccj, wc1000cj, wc200cj
Regeneration Solution ID	180501699012
Sufficient Volume for Batch QC	yes

Basis	Basis Description
T	Total/NA

The pound sign (#) in the amount added field denotes that the reagent was used undiluted. All calculations are performed using the stated concentration for this reagent.

9056A

Page 2 of 2

TestAmerica Denver  
Target Compound Quantitation Report

Data File: \\ChromNA\Denver\ChromData\WC\_IonChrom7\20180605-70676.b\02.0000.d  
 Lims ID: std L1  
 Client ID:  
 Sample Type: IC Calib Level: 1  
 Inject. Date: 05-Jun-2018 10:33:00 ALS Bottle#: 0 Worklist Smp#: 2  
 Injection Vol: 25.0 ul Dil. Factor: 1.0000  
 Sample Info: 280-0070676-002  
 Operator ID: Instrument ID: WC\_IonChrom7  
 Sublist: chrom-Anions\_IC7\*sub1  
 Method: \\ChromNA\Denver\ChromData\WC\_IonChrom7\20180605-70676.b\Anions\_IC7.m  
 Limit Group: Wet - Anions  
 Last Update: 05-Jun-2018 12:36:27 Calib Date: 05-Jun-2018 12:02:00  
 Integrator: Falcon  
 Quant Method: External Standard Quant By: Initial Calibration  
 Last ICal File: \\ChromNA\Denver\ChromData\WC\_IonChrom7\20180605-70676.b\07.0000.d  
 Column 1 : Det: 0005  
 Process Host: CTX0302

First Level Reviewer: jewellc Date: 05-Jun-2018 12:34:11

Compound	RT (min.)	Exp RT (min.)	Dlt RT (min.)	Response	Cal Amt ug/ml	OnCol Amt ug/ml	Flags
1 Fluoride	2.000	2.017	-0.017	6350951	0.2000	0.1911	
2 Chloride	3.233	3.208	0.025	17204213	1.00	0.9821	
3 Nitrite as N	3.725	3.717	0.008	8547902	0.2000	0.1933	
4 Bromide	5.917	5.808	0.109	1282303	0.2000	0.2100	
5 Nitrate as N	6.467	6.200	0.267	8707426	0.2000	0.2019	
6 Sulfate	9.300	8.850	0.450	14285710	1.00	0.9474	
7 Orthophosphate as P		11.492			ND	ND	

### QC Flag Legend

#### Processing Flags

ND - Not Detected or Marked ND

#### Reagents:


IC Cal low_00373	Amount Added: 0.02	Units: mL
IC CAL cl/so4_00202	Amount Added: 0.02	Units: mL

Report Date: 05-Jun-2018 12:36:27

Chrom Revision: 2.2 11-May-2018 08:54:46

TestAmerica Denver

Data File: \\ChromNA\\Denver\\ChromData\\WC\_IonChrom7\\20180605-70676.b\\02.0000.d  
Injection Date: 05-Jun-2018 10:33:00 Instrument ID: WC\_IonChrom7  
Lims ID: std L1 Operator ID:  
Client ID:  
Injection Vol: 25.0 ul ALS Bottle#: 0  
Method: Anions\_IC7 Dil. Factor: 1.0000  
Limit Group: Wet - Anions

Operator ID:  
Worklist Smp#: 2

TestAmerica Denver  
Target Compound Quantitation Report

Data File: \\ChromNA\Denver\ChromData\WC\_IonChrom7\20180605-70676.b\03.0000.d  
 Lims ID: std L2  
 Client ID:  
 Sample Type: IC Calib Level: 2  
 Inject. Date: 05-Jun-2018 10:51:00 ALS Bottle#: 0 Worklist Smp#: 3  
 Injection Vol: 25.0 ul Dil. Factor: 1.0000  
 Sample Info: 280-0070676-003  
 Operator ID: Instrument ID: WC\_IonChrom7  
 Sublist: chrom-Anions\_IC7\*sub1  
 Method: \\ChromNA\Denver\ChromData\WC\_IonChrom7\20180605-70676.b\Anions\_IC7.m  
 Limit Group: Wet - Anions  
 Last Update: 05-Jun-2018 12:36:27 Calib Date: 05-Jun-2018 12:02:00  
 Integrator: Falcon  
 Quant Method: External Standard Quant By: Initial Calibration  
 Last ICAL File: \\ChromNA\Denver\ChromData\WC\_IonChrom7\20180605-70676.b\07.0000.d  
 Column 1 : Det: 0005  
 Process Host: CTX0302

First Level Reviewer: jewellc Date: 05-Jun-2018 12:35:00

Compound	RT (min.)	Exp RT (min.)	Dlt RT (min.)	Response	Cal Amt ug/ml	OnCol Amt ug/ml	Flags
1 Fluoride	2.000	2.017	-0.017	14729157	0.5000	0.4955	
2 Chloride	3.225	3.208	0.017	42893961	2.50	2.49	
3 Nitrite as N	3.725	3.717	0.008	21664939	0.5000	0.5073	
4 Bromide	5.908	5.808	0.100	3452047	0.5000	0.4958	
5 Nitrate as N	6.450	6.200	0.250	22099291	0.5000	0.5037	
6 Sulfate	9.292	8.850	0.442	34173739	2.50	2.52	
7 Orthophosphate as P		11.492			ND	ND	

### QC Flag Legend

#### Processing Flags

ND - Not Detected or Marked ND

#### Reagents:

IC Cal low_00373	Amount Added: 0.05	Units: mL
IC CAL cl/so4_00202	Amount Added: 0.05	Units: mL


Report Date: 05-Jun-2018 12:36:27

Chrom Revision: 2.2 11-May-2018 08:54:46

TestAmerica Denver

Data File: \\ChromNA\\Denver\\ChromData\\WC\_IonChrom7\\20180605-70676.b\\03.0000.d  
Injection Date: 05-Jun-2018 10:51:00 Instrument ID: WC\_IonChrom7  
Lims ID: std L2 Operator ID:  
Client ID:  
Injection Vol: 25.0 ul ALS Bottle#: 0  
Method: Anions\_IC7 Dil. Factor: 1.0000  
Limit Group: Wet - Anions

Worklist Smp#: 3



TestAmerica Denver  
Target Compound Quantitation Report

Data File: \\ChromNA\Denver\ChromData\WC\_IonChrom7\20180605-70676.b\04.0000.d  
 Lims ID: std L3  
 Client ID:  
 Sample Type: IC Calib Level: 3  
 Inject. Date: 05-Jun-2018 11:09:00 ALS Bottle#: 0 Worklist Smp#: 4  
 Injection Vol: 25.0 ul Dil. Factor: 1.0000  
 Sample Info: 280-0070676-004  
 Misc. Info.: 4 F  
 Operator ID: Instrument ID: WC\_IonChrom7  
 Sublist: chrom-Anions\_IC7\*sub1  
 Method: \\ChromNA\Denver\ChromData\WC\_IonChrom7\20180605-70676.b\Anions\_IC7.m  
 Limit Group: Wet - Anions  
 Last Update: 05-Jun-2018 12:36:28 Calib Date: 05-Jun-2018 12:02:00  
 Integrator: Falcon  
 Quant Method: External Standard Quant By: Initial Calibration  
 Last ICal File: \\ChromNA\Denver\ChromData\WC\_IonChrom7\20180605-70676.b\07.0000.d  
 Column 1 : Det: 0005  
 Process Host: CTX0302

First Level Reviewer: jewelc Date: 05-Jun-2018 12:35:12

Compound	RT (min.)	Exp RT (min.)	Dlt RT (min.)	Response	Cal Amt ug/ml	OnCol Amt ug/ml	Flags
1 Fluoride	2.008	2.017	-0.009	29570255	1.00	1.03	
2 Chloride	3.225	3.208	0.017	86729334	5.00	5.07	
3 Nitrite as N	3.725	3.717	0.008	42498019	1.00	1.01	
4 Bromide	5.908	5.808	0.100	6969044	1.00	0.9592	
5 Nitrate as N	6.425	6.200	0.225	43441006	1.00	0.9845	
6 Sulfate	9.275	8.850	0.425	67795829	5.00	5.17	
7 Orthophosphate as P	12.025	11.492	0.533	12515848	1.00	1.01	

**Reagents:**

IC Cal low_00373	Amount Added: 0.10	Units: mL
IC CAL cl/so4_00202	Amount Added: 0.10	Units: mL


Report Date: 05-Jun-2018 12:36:28

Chrom Revision: 2.2 11-May-2018 08:54:46

TestAmerica Denver

Data File: \\ChromNA\\Denver\\ChromData\\WC\_IonChrom7\\20180605-70676.b\\04.0000.d  
Injection Date: 05-Jun-2018 11:09:00 Instrument ID: WC\_IonChrom7  
Lims ID: std L3 Operator ID:  
Client ID:  
Injection Vol: 25.0 ul ALS Bottle#: 0  
Method: Anions\_IC7 Dil. Factor: 1.0000  
Limit Group: Wet - Anions

Worklist Smp#: 4



TestAmerica Denver  
Target Compound Quantitation Report

Data File: \\ChromNA\Denver\ChromData\WC\_IonChrom7\20180605-70676.b\05.0000.d  
 Lims ID: std L4  
 Client ID:  
 Sample Type: IC Calib Level: 4  
 Inject. Date: 05-Jun-2018 11:27:00 ALS Bottle#: 0 Worklist Smp#: 5  
 Injection Vol: 25.0 ul Dil. Factor: 1.0000  
 Sample Info: 280-0070676-005  
 Misc. Info.: 5 F  
 Operator ID: Instrument ID: WC\_IonChrom7  
 Sublist: chrom-Anions\_IC7\*sub1  
 Method: \\ChromNA\Denver\ChromData\WC\_IonChrom7\20180605-70676.b\Anions\_IC7.m  
 Limit Group: Wet - Anions  
 Last Update: 05-Jun-2018 12:36:28 Calib Date: 05-Jun-2018 12:02:00  
 Integrator: Falcon  
 Quant Method: External Standard Quant By: Initial Calibration  
 Last ICal File: \\ChromNA\Denver\ChromData\WC\_IonChrom7\20180605-70676.b\07.0000.d  
 Column 1 : Det: 0005  
 Process Host: CTX0302

First Level Reviewer: jewelc Date: 05-Jun-2018 12:35:23

Compound	RT (min.)	Exp RT (min.)	Dlt RT (min.)	Response	Cal Amt ug/ml	OnCol Amt ug/ml	Flags
1 Fluoride	2.008	2.008	0.000	114940420	4.00	4.14	
2 Chloride	3.208	3.208	0.000	1031096518	60.0	60.6	
3 Nitrite as N	3.725	3.725	0.000	171553985	4.00	4.10	
4 Bromide	5.875	5.875	0.000	29791673	4.00	3.97	
5 Nitrate as N	6.342	6.342	0.000	176382525	4.00	3.98	
6 Sulfate	9.083	9.083	0.000	772995817	60.0	60.9	
7 Orthophosphate as P	11.667	11.667	0.000	65509655	4.00	3.94	

**Reagents:**

IC Cal low_00373	Amount Added: 0.40	Units: mL
IC CAL cl/so4_00202	Amount Added: 1.20	Units: mL


Report Date: 05-Jun-2018 12:36:28

Chrom Revision: 2.2 11-May-2018 08:54:46

TestAmerica Denver

Data File: \\ChromNA\\Denver\\ChromData\\WC\_IonChrom7\\20180605-70676.b\\05.0000.d  
Injection Date: 05-Jun-2018 11:27:00 Instrument ID: WC\_IonChrom7  
Lims ID: std L4 Operator ID:  
Client ID:  
Injection Vol: 25.0 ul Worklist Smp#: 5  
Method: Anions\_IC7 Dil. Factor: 1.0000  
Limit Group: Wet - Anions

05.0000, 05.0000.d



TestAmerica Denver  
Target Compound Quantitation Report

Data File: \\ChromNA\Denver\ChromData\WC\_IonChrom7\20180605-70676.b\06.0000.d  
 Lims ID: std L5  
 Client ID:  
 Sample Type: IC Calib Level: 5  
 Inject. Date: 05-Jun-2018 11:44:00 ALS Bottle#: 0 Worklist Smp#: 6  
 Injection Vol: 25.0 ul Dil. Factor: 1.0000  
 Sample Info: 280-0070676-006  
 Misc. Info.: 6 F  
 Operator ID: Instrument ID: WC\_IonChrom7  
 Sublist: chrom-Anions\_IC7\*sub1  
 Method: \\ChromNA\Denver\ChromData\WC\_IonChrom7\20180605-70676.b\Anions\_IC7.m  
 Limit Group: Wet - Anions  
 Last Update: 05-Jun-2018 12:36:26 Calib Date: 05-Jun-2018 12:02:00  
 Integrator: Falcon  
 Quant Method: External Standard Quant By: Initial Calibration  
 Last ICal File: \\ChromNA\Denver\ChromData\WC\_IonChrom7\20180605-70676.b\07.0000.d  
 Column 1 : Det: 0005  
 Process Host: CTX0302

First Level Reviewer: jewelc Date: 05-Jun-2018 12:36:26

Compound	RT (min.)	Exp RT (min.)	Dlt RT (min.)	Response	Cal Amt ug/ml	OnCol Amt ug/ml	Flags
1 Fluoride	2.017	2.017	0.000	221467033	8.00	8.01	
2 Chloride	3.208	3.208	0.000	2047237442	120.0	120.3	
3 Nitrite as N	3.725	3.717	0.008	333142558	8.00	7.96	
4 Bromide	5.833	5.808	0.025	60688987	8.00	8.04	
5 Nitrate as N	6.250	6.200	0.050	355903515	8.00	8.02	
6 Sulfate	8.967	8.850	0.117	1523157981	120.0	120.2	
7 Orthophosphate as P	11.550	11.492	0.058	139065974	8.00	8.00	

**Reagents:**

IC Cal low_00373	Amount Added: 0.80	Units: mL
IC CAL cl/so4_00202	Amount Added: 2.40	Units: mL


Report Date: 05-Jun-2018 12:36:26

Chrom Revision: 2.2 11-May-2018 08:54:46

TestAmerica Denver

Data File: \\ChromNA\\Denver\\ChromData\\WC\_IonChrom7\\20180605-70676.b\\06.0000.d  
Injection Date: 05-Jun-2018 11:44:00 Instrument ID: WC\_IonChrom7  
Lims ID: std L5 Operator ID:  
Client ID:  
Injection Vol: 25.0 ul ALS Bottle#: 0  
Method: Anions\_IC7 Dil. Factor: 1.0000  
Limit Group: Wet - Anions

Worklist Smp#: 6



TestAmerica Denver  
Target Compound Quantitation Report

Data File: \\ChromNA\Denver\ChromData\WC\_IonChrom7\20180605-70676.b\07.0000.d  
 Lims ID: std L6  
 Client ID:  
 Sample Type: IC Calib Level: 6  
 Inject. Date: 05-Jun-2018 12:02:00 ALS Bottle#: 0 Worklist Smp#: 7  
 Injection Vol: 25.0 ul Dil. Factor: 1.0000  
 Sample Info: 280-0070676-007  
 Misc. Info.: 7 F  
 Operator ID: Instrument ID: WC\_IonChrom7  
 Sublist: chrom-Anions\_IC7\*sub1  
 Method: \\ChromNA\Denver\ChromData\WC\_IonChrom7\20180605-70676.b\Anions\_IC7.m  
 Limit Group: Wet - Anions  
 Last Update: 05-Jun-2018 12:36:29 Calib Date: 05-Jun-2018 12:02:00  
 Integrator: Falcon  
 Quant Method: External Standard Quant By: Initial Calibration  
 Last ICal File: \\ChromNA\Denver\ChromData\WC\_IonChrom7\20180605-70676.b\07.0000.d  
 Column 1 : Det: 0005  
 Process Host: CTX0302

First Level Reviewer: jewelc Date: 05-Jun-2018 12:35:56

Compound	RT (min.)	Exp RT (min.)	Dlt RT (min.)	Response	Cal Amt ug/ml	OnCol Amt ug/ml	Flags
1 Fluoride	2.017	2.008	0.009	271729025	10.0	9.83	
2 Chloride	3.208	3.208	0.000	3388060477	200.0	199.1	
3 Nitrite as N	3.717	3.725	-0.008	415327633	10.0	9.93	
4 Bromide	5.808	5.875	-0.067	75830729	10.0	10.0	
5 Nitrate as N	6.200	6.342	-0.142	443857326	10.0	10.0	
6 Sulfate	8.850	9.083	-0.233	2518985817	200.0	198.8	
7 Orthophosphate as P	11.492	11.667	-0.175	176332183	10.0	10.1	

**Reagents:**

IC Cal low_00373	Amount Added: 1.00	Units: mL
IC CAL cl/so4_00202	Amount Added: 4.00	Units: mL


Report Date: 05-Jun-2018 12:36:29

Chrom Revision: 2.2 11-May-2018 08:54:46

## TestAmerica Denver

Data File: \\ChromNA\\Denver\\ChromData\\WC\_IonChrom7\\20180605-70676.b\\07.0000.d  
Injection Date: 05-Jun-2018 12:02:00 Instrument ID: WC\_IonChrom7  
Lims ID: std L6 Operator ID:  
Client ID:  
Injection Vol: 25.0 ul Worklist Smp#: 7  
Method: Anions\_IC7 Dil. Factor: 1.0000  
Limit Group: Wet - Anions

07.0000, 07.0000.d



419404 / 419405

## **IC Instrument Information**

WL: 70676 Inst ID: 7 Analysis Date: 06/05/18 Analyst: JP

### Dilutions

Phimix  
b19/19

R:\QAIEdit\FORMSIGCIC Instrument Information Rev. 1.doc  
Rev. 1, 8/11/2014

Page 1 of 1

## IC Instrument Information

WL: 70676 Inst ID: 7 Analysis Date: 06/05/18 Analyst: T

Rush	Job No.	Samples	Anions	QC Req	HT Exp
<input type="checkbox"/>	<u>110516</u>	<u>5</u>	F Cl NO <sub>2</sub> Br NO <sub>3</sub> PO <sub>4</sub> SO <sub>4</sub>	MS/D	1 <u>Σ 5</u>
<input checked="" type="checkbox"/>	<u>103735</u>	<u>1</u>	F Cl NO <sub>2</sub> Br NO <sub>3</sub> PO <sub>4</sub> SO <sub>4</sub>	MS/D	
<input type="checkbox"/>	<u>110464</u>	<u>1</u>	F Cl NO <sub>2</sub> Br NO <sub>3</sub> PO <sub>4</sub> SO <sub>4</sub>	MS/D	
<input type="checkbox"/>	<u>110474</u>	<u>2</u>	F Cl NO <sub>2</sub> Br NO <sub>3</sub> PO <sub>4</sub> SO <sub>4</sub>	MS/D	
<input type="checkbox"/>	<u>109875</u>	<u>7</u>	F Cl NO <sub>2</sub> Br NO <sub>3</sub> PO <sub>4</sub> SO <sub>4</sub>	MS/D	
<input type="checkbox"/>	<u>109813</u>	<u>1</u>	F Cl NO <sub>2</sub> Br NO <sub>3</sub> PO <sub>4</sub> SO <sub>4</sub>	MS/D	
<input type="checkbox"/>	<u>109962</u>	<u>7</u>	F Cl NO <sub>2</sub> Br NO <sub>3</sub> PO <sub>4</sub> SO <sub>4</sub>	MS/D	
<input type="checkbox"/>			F Cl NO <sub>2</sub> Br NO <sub>3</sub> PO <sub>4</sub> SO <sub>4</sub>	MS/D	
<input type="checkbox"/>			F Cl NO <sub>2</sub> Br NO <sub>3</sub> PO <sub>4</sub> SO <sub>4</sub>	MS/D	
<input type="checkbox"/>			F Cl NO <sub>2</sub> Br NO <sub>3</sub> PO <sub>4</sub> SO <sub>4</sub>	MS/D	
<input type="checkbox"/>			F Cl NO <sub>2</sub> Br NO <sub>3</sub> PO <sub>4</sub> SO <sub>4</sub>	MS/D	
<input type="checkbox"/>			F Cl NO <sub>2</sub> Br NO <sub>3</sub> PO <sub>4</sub> SO <sub>4</sub>	MS/D	
<input type="checkbox"/>			F Cl NO <sub>2</sub> Br NO <sub>3</sub> PO <sub>4</sub> SO <sub>4</sub>	MS/D	
<input type="checkbox"/>			F Cl NO <sub>2</sub> Br NO <sub>3</sub> PO <sub>4</sub> SO <sub>4</sub>	MS/D	
<input type="checkbox"/>			F Cl NO <sub>2</sub> Br NO <sub>3</sub> PO <sub>4</sub> SO <sub>4</sub>	MS/D	

### Dilutions

Job No.	Samples	Anions	Dilution	Reason
		F Cl NO <sub>2</sub> Br NO <sub>3</sub> PO <sub>4</sub> SO <sub>4</sub>		
		F Cl NO <sub>2</sub> Br NO <sub>3</sub> PO <sub>4</sub> SO <sub>4</sub>		
		F Cl NO <sub>2</sub> Br NO <sub>3</sub> PO <sub>4</sub> SO <sub>4</sub>		
		F Cl NO <sub>2</sub> Br NO <sub>3</sub> PO <sub>4</sub> SO <sub>4</sub>		
		F Cl NO <sub>2</sub> Br NO <sub>3</sub> PO <sub>4</sub> SO <sub>4</sub>		
		F Cl NO <sub>2</sub> Br NO <sub>3</sub> PO <sub>4</sub> SO <sub>4</sub>		
		F Cl NO <sub>2</sub> Br NO <sub>3</sub> PO <sub>4</sub> SO <sub>4</sub>		
		F Cl NO <sub>2</sub> Br NO <sub>3</sub> PO <sub>4</sub> SO <sub>4</sub>		
		F Cl NO <sub>2</sub> Br NO <sub>3</sub> PO <sub>4</sub> SO <sub>4</sub>		
		F Cl NO <sub>2</sub> Br NO <sub>3</sub> PO <sub>4</sub> SO <sub>4</sub>		

**TestAmerica Laboratories**  
**Initial Calibration Summary Report**

Method: \\ChromNA\\Denver\\ChromData\\WC\_IonChrom7\\20180605-70676.b\\Anions\_IC7.m

Instrument: WC\_IonChrom7

Lims Location: 280

Lock State: Initial Calib Locked

Cpnd Order: Retention Time

Integrator: Falcon

Last Modified: 05-Jun-2018 12:36:26

No.Compounds:7

**Initial Calibration Batches**

Ical Batch: \\ChromNA\\Denver\\ChromData\\WC\_IonChrom7\\20180605-70676.b

Inj Date : 05-Jun-2018 10:33:00, Sublist: chrom-Anions\_IC7\*sub1

Detector 1: 0005

Compound	Wet - Anions				Wet - Anions 28D			
	D	M1	M2	Err	D	M1	M2	Err
1 Fluoride	1091607	275205E		1.000	1091607	275205E		1.000
2 Chloride	493398	1701483		1.000	493398	1701483		1.000
3 Nitrite as N	475295	4176722		1.000	475295	4176722		1.000
4 Bromide	-311280	7589977		1.000	-311280	7589977		1.000
5 Nitrate as N	-255286	4438492		1.000	-255286	4438492		1.000
6 Sulfate	2292933	1265801		1.000	2292933	1265801		1.000
7 Orthophosphate as P	-578917	1811218	R2, R4	1.000*	-578917	1811218	R2, R4	1.000*

ICalib Error Legend

R2, Missing the Required Number of Calibration Points

R4, Curve Zero Intercept Is > Reporting Limit

Phuvinh  
6/17/18

TestAmerica Laboratories  
Initial Calibration Summary Report

Method: \\ChromNA\\Denver\\ChromData\\WC\_IonChrom7\\20180605-70676.b\\Anions\_IC7.m  
 Instrument: WC\_IonChrom7 Lims Location: 280  
 Lock State: Initial Calib Locked Cpnd Order: Retention Time  
 Integrator: Falcon Last Modified: 05-Jun-2018 12:36:26  
 No.Compounds:7

**Initial Calibration Batches**

Ical Batch: \\ChromNA\\Denver\\ChromData\\WC\_IonChrom7\\20180605-70676.b  
 Inj Date : 05-Jun-2018 10:33:00, Sublist: chrom-Anions\_IC7\*sub1

Detector 1: 0005

Compound	Wet - Anions				Wet - Anions 28D			
	b	M1	M2	Err	b	M1	M2	Err
1 Fluoride	1091607	2752055		1.000	1091607	2752055		1.000
2 Chloride	493398	1701483		1.000	493398	1701483		1.000
3 Nitrite as N	475295	4176722		1.000	475295	4176722		1.000
4 Bromide	-311280	7589977		1.000	-311280	7589977		1.000
5 Nitrate as N	-255286	4438492		1.000	-255286	4438492		1.000
6 Sulfate	2292933	1265801		1.000	2292933	1265801		1.000
7 Orthophosphate as P	-578917	1811218	R2, R4	1.000*	-578917	1811218	R2, R4	1.000*

**ICalib Error Legend**

R2, Missing the Required Number of Calibration Points

R4, Curve Zero Intercept is > Reporting Limit

TestAmerica Denver  
Target Compound Quantitation Report

Data File: \\ChromNA\Denver\ChromData\WC\_IonChrom7\20180605-70676.b\08.0000.d  
 Lims ID: icv  
 Client ID:  
 Sample Type: ICV  
 Inject. Date: 05-Jun-2018 12:20:00 ALS Bottle#: 0 Worklist Smp#: 8  
 Injection Vol: 25.0 ul Dil. Factor: 1.0000  
 Sample Info: 280-0070676-008  
 Misc. Info.: 8 F  
 Operator ID: Instrument ID: WC\_IonChrom7  
 Sublist:  
 Method: \\ChromNA\Denver\ChromData\WC\_IonChrom7\20180605-70676.b\Anions\_IC7.m  
 Limit Group: Wet - Anions  
 Last Update: 06-Jun-2018 10:01:58 Calib Date: 05-Jun-2018 12:02:00  
 Integrator: Falcon  
 Quant Method: External Standard Quant By: Initial Calibration  
 Last ICal File: \\ChromNA\Denver\ChromData\WC\_IonChrom7\20180605-70676.b\07.0000.d

Column 1 : Det: 0005  
 Process Host: CTX0303

Compound	RT (min.)	Exp RT (min.)	Dlt RT (min.)	Response	Cal Amt ug/ml	OnCol Amt ug/ml	Flags
1 Fluoride	2.008	2.008	0.000	115146856	4.00	4.14	
2 Chloride	3.208	3.208	0.000	1399306524	80.0	82.2	
3 Nitrite as N	3.717	3.717	0.000	173786735	4.00	4.15	
4 Bromide	5.867	5.867	0.000	30184915	4.00	4.02	
5 Nitrate as N	6.333	6.333	0.000	181727518	4.00	4.10	
6 Sulfate	9.033	9.033	0.000	1045532426	80.0	82.4	
7 Orthophosphate as P	11.658	11.658	0.000	68190758	4.00	4.08	

**Reagents:**

IC CL ICV_00014	Amount Added: 0.40	Units: mL
IC ICV 5_00201	Amount Added: 0.40	Units: mL
IC SO4 ICV_00017	Amount Added: 0.40	Units: mL


Report Date: 06-Jun-2018 10:01:59

Chrom Revision: 2.2 11-May-2018 08:54:46

TestAmerica Denver

Data File: \\ChromNA\\Denver\\ChromData\\WC\_IonChrom7\\20180605-70676.b\\08.0000.d  
Injection Date: 05-Jun-2018 12:20:00 Instrument ID: WC\_IonChrom7  
Lims ID: icv Operator ID:  
Client ID:  
Injection Vol: 25.0 ul ALS Bottle#: 0  
Method: Anions\_IC7 Dil. Factor: 1.0000  
Limit Group: Wet - Anions

Worklist Smp#: 8



TestAmerica Denver  
Target Compound Quantitation Report

Data File: \\ChromNA\Denver\ChromData\WC\_IonChrom7\20180605-70676.b\09.0000.d  
 Lims ID: icb  
 Client ID:  
 Sample Type: ICB  
 Inject. Date: 05-Jun-2018 12:38:00 ALS Bottle#: 0 Worklist Smp#: 9  
 Injection Vol: 25.0 ul Dil. Factor: 1.0000  
 Sample Info: 280-0070676-009  
 Misc. Info.: 9 F  
 Operator ID: Instrument ID: WC\_IonChrom7  
 Method: \\ChromNA\Denver\ChromData\WC\_IonChrom7\20180605-70676.b\Anions\_IC7.m  
 Limit Group: Wet - Anions  
 Last Update: 06-Jun-2018 10:01:58 Calib Date: 05-Jun-2018 12:02:00  
 Integrator: Falcon  
 Quant Method: External Standard Quant By: Initial Calibration  
 Last ICal File: \\ChromNA\Denver\ChromData\WC\_IonChrom7\20180605-70676.b\07.0000.d  
 Column 1 : Det: 0005  
 Process Host: CTX0303

Compound	RT (min.)	Exp RT (min.)	Dlt RT (min.)	Response	Cal Amt ug/ml	OnCol Amt ug/ml	Flags
1 Fluoride	1.992	2.008	-0.016	586550	-0.0184		
2 Chloride	3.217	3.208	0.009	186324	-0.0180		
3 Nitrite as N		3.717			ND		
4 Bromide		5.867			ND		
5 Nitrate as N	6.467	6.333	0.134	57287	0.007042		
6 Sulfate	9.308	9.033	0.275	538489	-0.1386		
7 Orthophosphate as P		11.658			ND		


Report Date: 06-Jun-2018 10:02:01

Chrom Revision: 2.2 11-May-2018 08:54:46

## TestAmerica Denver

Data File: \\ChromNA\\Denver\\ChromData\\WC\_IonChrom7\\20180605-70676.b\\09.0000.d  
Injection Date: 05-Jun-2018 12:38:00 Instrument ID: WC\_IonChrom7  
Lims ID: icb Operator ID:  
Client ID:  
Injection Vol: 25.0 ul ALS Bottle#: 0  
Method: Anions\_IC7 Dil. Factor: 1.0000  
Limit Group: Wet - Anions

Worklist Smp#: 9



**IC Instrument Information**

422998 / 422999

WL: 72165 Inst ID: 7 Analysis Date: 07/20/18 Analyst: CJ

Rush	Job No.	Samples	Anions	QC Req	HT Exp
<input type="checkbox"/>	<u>112132</u>	<u>2</u>	F Cl NO <sub>2</sub> Br NO <sub>3</sub> PO <sub>4</sub> SO <sub>4</sub>	MS/D	<u>2</u>
<input type="checkbox"/>	<u>112220</u>	<u>4</u>	F Cl NO <sub>2</sub> Br NO <sub>3</sub> PO <sub>4</sub> SO <sub>4</sub>	MS/D	
<input type="checkbox"/>	<u>11224</u>	<u>5</u>	F Cl NO <sub>2</sub> Br NO <sub>3</sub> PO <sub>4</sub> SO <sub>4</sub>	MS/D	
<input type="checkbox"/>	<u>112225</u>	<u>4</u>	F Cl NO <sub>2</sub> Br NO <sub>3</sub> PO <sub>4</sub> SO <sub>4</sub>	MS/D	
<input type="checkbox"/>	<u>111519</u> x	<u>2</u>	F Cl NO <sub>2</sub> Br NO <sub>3</sub> PO <sub>4</sub> SO <sub>4</sub>	MS/D	<u>1</u>
<input type="checkbox"/>	<u>111469</u>	<u>1</u>	F Cl NO <sub>2</sub> Br NO <sub>3</sub> PO <sub>4</sub> SO <sub>4</sub>	MS/D	<u>07/20/18</u>
<input type="checkbox"/>	<u>111554</u>	<u>3</u>	F Cl NO <sub>2</sub> Br NO <sub>3</sub> PO <sub>4</sub> SO <sub>4</sub>	MS/D	
<input type="checkbox"/>			F Cl NO <sub>2</sub> Br NO <sub>3</sub> PO <sub>4</sub> SO <sub>4</sub>	MS/D	
<input type="checkbox"/>			F Cl NO <sub>2</sub> Br NO <sub>3</sub> PO <sub>4</sub> SO <sub>4</sub>	MS/D	
<input type="checkbox"/>			F Cl NO <sub>2</sub> Br NO <sub>3</sub> PO <sub>4</sub> SO <sub>4</sub>	MS/D	
<input type="checkbox"/>			F Cl NO <sub>2</sub> Br NO <sub>3</sub> PO <sub>4</sub> SO <sub>4</sub>	MS/D	
<input type="checkbox"/>			F Cl NO <sub>2</sub> Br NO <sub>3</sub> PO <sub>4</sub> SO <sub>4</sub>	MS/D	
<input type="checkbox"/>			F Cl NO <sub>2</sub> Br NO <sub>3</sub> PO <sub>4</sub> SO <sub>4</sub>	MS/D	
<input type="checkbox"/>			F Cl NO <sub>2</sub> Br NO <sub>3</sub> PO <sub>4</sub> SO <sub>4</sub>	MS/D	

**Dilutions**

Job No.	Samples	Anions	Dilution	Reason
		F Cl NO <sub>2</sub> Br NO <sub>3</sub> PO <sub>4</sub> SO <sub>4</sub>		
		F Cl NO <sub>2</sub> Br NO <sub>3</sub> PO <sub>4</sub> SO <sub>4</sub>		
		F Cl NO <sub>2</sub> Br NO <sub>3</sub> PO <sub>4</sub> SO <sub>4</sub>		
		F Cl NO <sub>2</sub> Br NO <sub>3</sub> PO <sub>4</sub> SO <sub>4</sub>		
		F Cl NO <sub>2</sub> Br NO <sub>3</sub> PO <sub>4</sub> SO <sub>4</sub>		
		F Cl NO <sub>2</sub> Br NO <sub>3</sub> PO <sub>4</sub> SO <sub>4</sub>		
		F Cl NO <sub>2</sub> Br NO <sub>3</sub> PO <sub>4</sub> SO <sub>4</sub>		
		F Cl NO <sub>2</sub> Br NO <sub>3</sub> PO <sub>4</sub> SO <sub>4</sub>		
		F Cl NO <sub>2</sub> Br NO <sub>3</sub> PO <sub>4</sub> SO <sub>4</sub>		
		F Cl NO <sub>2</sub> Br NO <sub>3</sub> PO <sub>4</sub> SO <sub>4</sub>		

## IC Instrument Information

WL: 72165 Inst ID: 7 Analysis Date: 07/20/18 Analyst: CJ

Rush	Job No.	Samples	Anions	QC Req	HT Exp
<input type="checkbox"/>	<u>112132</u>	<u>2</u>	F Cl NO <sub>2</sub> Br NO <sub>3</sub> PO <sub>4</sub> SO <sub>4</sub>	MS/D	<u>2</u>
<input type="checkbox"/>	<u>112220</u>	<u>4</u>	F Cl NO <sub>2</sub> Br NO <sub>3</sub> PO <sub>4</sub> SO <sub>4</sub>	MS/D	
<input type="checkbox"/>	<u>112214</u>	<u>5</u>	F Cl NO <sub>2</sub> Br NO <sub>3</sub> PO <sub>4</sub> SO <sub>4</sub>	MS/D	
<input type="checkbox"/>	<u>112225</u>	<u>4</u>	F Cl NO <sub>2</sub> Br NO <sub>3</sub> PO <sub>4</sub> SO <sub>4</sub>	MS/D	
<input type="checkbox"/>	<u>111519</u>	<u>2</u>	F Cl NO <sub>2</sub> Br NO <sub>3</sub> PO <sub>4</sub> SO <sub>4</sub>	MS/D	
<input type="checkbox"/>	<u>111468</u>	<u>1</u>	F Cl NO <sub>2</sub> Br NO <sub>3</sub> PO <sub>4</sub> SO <sub>4</sub>	MS/D	<u>07/20/18</u>
<input type="checkbox"/>	<u>111554</u>	<u>3</u>	F Cl NO <sub>2</sub> Br NO <sub>3</sub> PO <sub>4</sub> SO <sub>4</sub>	MS/D	
<input type="checkbox"/>			F Cl NO <sub>2</sub> Br NO <sub>3</sub> PO <sub>4</sub> SO <sub>4</sub>	MS/D	
<input type="checkbox"/>			F Cl NO <sub>2</sub> Br NO <sub>3</sub> PO <sub>4</sub> SO <sub>4</sub>	MS/D	
<input type="checkbox"/>			F Cl NO <sub>2</sub> Br NO <sub>3</sub> PO <sub>4</sub> SO <sub>4</sub>	MS/D	
<input type="checkbox"/>			F Cl NO <sub>2</sub> Br NO <sub>3</sub> PO <sub>4</sub> SO <sub>4</sub>	MS/D	
<input type="checkbox"/>			F Cl NO <sub>2</sub> Br NO <sub>3</sub> PO <sub>4</sub> SO <sub>4</sub>	MS/D	
<input type="checkbox"/>			F Cl NO <sub>2</sub> Br NO <sub>3</sub> PO <sub>4</sub> SO <sub>4</sub>	MS/D	
<input type="checkbox"/>			F Cl NO <sub>2</sub> Br NO <sub>3</sub> PO <sub>4</sub> SO <sub>4</sub>	MS/D	
<input type="checkbox"/>			F Cl NO <sub>2</sub> Br NO <sub>3</sub> PO <sub>4</sub> SO <sub>4</sub>	MS/D	

### Dilutions

Job No.	Samples	Anions	Dilution	Reason
		F Cl NO <sub>2</sub> Br NO <sub>3</sub> PO <sub>4</sub> SO <sub>4</sub>		
		F Cl NO <sub>2</sub> Br NO <sub>3</sub> PO <sub>4</sub> SO <sub>4</sub>		
		F Cl NO <sub>2</sub> Br NO <sub>3</sub> PO <sub>4</sub> SO <sub>4</sub>		
		F Cl NO <sub>2</sub> Br NO <sub>3</sub> PO <sub>4</sub> SO <sub>4</sub>		
		F Cl NO <sub>2</sub> Br NO <sub>3</sub> PO <sub>4</sub> SO <sub>4</sub>		
		F Cl NO <sub>2</sub> Br NO <sub>3</sub> PO <sub>4</sub> SO <sub>4</sub>		
		F Cl NO <sub>2</sub> Br NO <sub>3</sub> PO <sub>4</sub> SO <sub>4</sub>		
		F Cl NO <sub>2</sub> Br NO <sub>3</sub> PO <sub>4</sub> SO <sub>4</sub>		
		F Cl NO <sub>2</sub> Br NO <sub>3</sub> PO <sub>4</sub> SO <sub>4</sub>		
		F Cl NO <sub>2</sub> Br NO <sub>3</sub> PO <sub>4</sub> SO <sub>4</sub>		
		F Cl NO <sub>2</sub> Br NO <sub>3</sub> PO <sub>4</sub> SO <sub>4</sub>		

TestAmerica Laboratories  
Initial Calibration Summary Report

Method: \\ChromNA\\Denver\\ChromData\\WC\_IonChrom7\\20180714-71975.b\\Anions\_IC7.m  
 Instrument: WC\_IonChrom7 Lims Location: 280  
 Lock State: Initial Calib Locked Cpnd Order: Retention Time  
 Integrator: Falcon Last Modified: 16-Jul-2018 23:38:47  
 No.Compounds:7

**Initial Calibration Batches**

Ical Batch: \\ChromNA\\Denver\\ChromData\\WC\_IonChrom7\\20180605-70676.b  
 Inj Date : 05-Jun-2018 10:33:00, Sublist: chrom-Anions\_IC7\*sub1

Detector 1: 0005

Compound	Wet - Anions				Wet - Anions 28D			
	b	M1	M2	Err	b	M1	M2	Err
1 Fluoride	1091607	2752055		1.000	1091607	2752055		1.000
2 Chloride	493398	1701483		1.000	493398	1701483		1.000
3 Nitrite as N	475295	4176722		1.000	475295	4176722		1.000
4 Bromide	-311280	7589977		1.000	-311280	7589977		1.000
5 Nitrate as N	-255286	4438492		1.000	-255286	4438492		1.000
6 Sulfate	2292932	1265801		1.000	2292933	1265801		1.000
7 Orthophosphate as P	-578917	1811218	R2, R4	1.000*	-578917	1811218	R2, R4	1.000*

ICalib Error Legend

R2, Missing the Required Number of Calibration Points

R4, Curve Zero Intercept is > Reporting Limit

TestAmerica Laboratories  
Initial Calibration Summary Report

Method: \\ChromNA\\Denver\\ChromData\\WC\_IonChrom7\\20180714-71975.b\\Anions\_IC7.m

Instrument: WC\_IonChrom7

Lims Location: 280

Lock State: Initial Calib Locked

Cpnd Order: Retention Time

Integrator: Falcon

Last Modified: 16-Jul-2018 23:38:47

No.Compounds:7

**Initial Calibration Batches**

Ical Batch: \\ChromNA\\Denver\\ChromData\\WC\_IonChrom7\\20180605-70676.b

Inj Date : 05-Jun-2018 10:33:00, Sublist: chrom-Anions\_IC7\*sub1

Detector 1: 0005

Compound	Wet - Anions				Wet - Anions 28D			
	b	M1	M2	Err	b	M1	M2	Err
1 Fluoride	1091607	2752055		1.000	1091607	2752055		1.000
2 Chloride	493398	1701483		1.000	493398	1701483		1.000
3 Nitrite as N	475295	4176722		1.000	475295	4176722		1.000
4 Bromide	-311280	7589977		1.000	-311280	7589977		1.000
5 Nitrate as N	-255286	4438492		1.000	-255286	4438492		1.000
6 Sulfate	2292932	1265801		1.000	2292932	1265801		1.000
7 Orthophosphate as P	-578917	1811218	R2, R4	1.000*	-578917	1811218	R2, R4	1.000*

ICalib Error Legend

R2, Missing the Required Number of Calibration Points

R4, Curve Zero Intercept is > Reporting Limit

Panida R.  
7/24/18

TestAmerica Denver  
Target Compound Quantitation Report

Data File: \\ChromNA\Denver\ChromData\WC\_IonChrom7\20180720-72165.b\01.0000.d  
 Lims ID: ccv  
 Client ID:  
 Sample Type: CCV  
 Inject. Date: 20-Jul-2018 10:14:00 ALS Bottle#: 0 Worklist Smp#: 1  
 Injection Vol: 25.0 ul Dil. Factor: 1.0000  
 Sample Info: 280-0072165-001  
 Operator ID: Instrument ID: WC\_IonChrom7  
 Sublist: chrom-Anions\_IC7\*sub1  
 Method: \\ChromNA\Denver\ChromData\WC\_IonChrom7\20180720-72165.b\Anions\_IC7.m  
 Limit Group: Wet - Anions  
 Last Update: 23-Jul-2018 09:24:33 Calib Date: 05-Jun-2018 12:02:00  
 Integrator: Falcon  
 Quant Method: External Standard Quant By: Initial Calibration  
 Last ICAL File: \\ChromNA\Denver\ChromData\WC\_IonChrom7\20180605-70676.b\07.0000.d  
 Column 1 : Det: 0005  
 Process Host: CTX0306

Compound	RT (min.)	Exp RT (min.)	Dlt RT (min.)	Response	Cal Amt ug/ml	OnCol Amt ug/ml	Flags
1 Fluoride	2.000	2.000	0.000	141928588	5.00	5.12	
2 Chloride	3.167	3.167	0.000	1666988912	100.0	97.9	
3 Nitrite as N	3.667	3.667	0.000	214207965	5.00	5.12	
4 Bromide	5.775	5.775	0.000	36476795	5.00	4.85	
5 Nitrate as N	6.217	6.217	0.000	222322777	5.00	5.01	
6 Sulfate	8.900	8.900	0.000	1315231902	100.0	103.7	
7 Orthophosphate as P	11.625	11.625	0.000	86871617	5.00	5.12	

**Reagents:**

IC LCS\_01288 Amount Added: 5.00 Units: mL

Report Date: 23-Jul-2018 09:24:34


Chrom Revision: 2.3 19-Jul-2018 15:14:50

TestAmerica Denver

Data File: \\ChromNA\\Denver\\ChromData\\WC\_IonChrom7\\20180720-72165.b\\01.0000.d  
Injection Date: 20-Jul-2018 10:14:00 Instrument ID: WC\_IonChrom7  
Lims ID: ccv Operator ID:  
Client ID:  
Injection Vol: 25.0 ul ALS Bottle#: 0  
Method: Anions\_IC7 Dil. Factor: 1.0000  
Limit Group: Wet - Anions

Worklist Smp#: 1

01.0000, 01.0000.d



TestAmerica Denver  
Target Compound Quantitation Report

Data File: \\ChromNA\Denver\ChromData\WC\_IonChrom7\20180720-72165.b\02.0000.d  
 Lims ID: ccb  
 Client ID:  
 Sample Type: CCB  
 Inject. Date: 20-Jul-2018 10:32:00      ALS Bottle#: 0      Worklist Smp#: 2  
 Injection Vol: 25.0 ul      Dil. Factor: 1.0000  
 Sample Info: 280-0072165-002  
 Operator ID:      Instrument ID: WC\_IonChrom7  
 Method: \\ChromNA\Denver\ChromData\WC\_IonChrom7\20180720-72165.b\Anions\_IC7.m  
 Limit Group: Wet - Anions  
 Last Update: 23-Jul-2018 09:24:33      Calib Date: 05-Jun-2018 12:02:00  
 Integrator: Falcon  
 Quant Method: External Standard      Quant By: Initial Calibration  
 Last ICal File: \\ChromNA\Denver\ChromData\WC\_IonChrom7\20180605-70676.b\07.0000.d  
 Column 1 :      Det: 0005  
 Process Host: CTX0306

Compound	RT (min.)	Exp RT (min.)	Dlt RT (min.)	Response	Cal Amt ug/ml	OnCol Amt ug/ml	Flags
1 Fluoride	1.975	2.000	-0.025	282638	-0.0294		
2 Chloride	3.175	3.158	0.017	556331	0.003699		
3 Nitrite as N		3.658			ND		
4 Bromide		5.742			ND		
5 Nitrate as N	6.375	6.183	0.192	104080	0.008097		
6 Sulfate	9.242	8.933	0.309	4655434	0.1866		
7 Orthophosphate as P		11.775			ND		


Report Date: 23-Jul-2018 09:24:36

Chrom Revision: 2.3 19-Jul-2018 15:14:50

## TestAmerica Denver

Data File: \\ChromNA\\Denver\\ChromData\\WC\_IonChrom7\\20180720-72165.b\\02.0000.d  
Injection Date: 20-Jul-2018 10:32:00 Instrument ID: WC\_IonChrom7  
Lims ID: ccb Operator ID:  
Client ID:  
Injection Vol: 25.0 ul ALS Bottle#: 0  
Method: Anions\_IC7 Dil. Factor: 1.0000  
Limit Group: Wet - Anions

Worklist Smp#: 2



TestAmerica Denver  
Target Compound Quantitation Report

Data File: \\ChromNA\Denver\ChromData\WC\_IonChrom7\20180720-72165.b\11.0000.d  
 Lims ID: lcs  
 Client ID:  
 Sample Type: LCS  
 Inject. Date: 20-Jul-2018 13:12:00 ALS Bottle#: 0 Worklist Smp#: 11  
 Injection Vol: 25.0 ul Dil. Factor: 1.0000  
 Sample Info: 280-0072165-011  
 Misc. Info.: 11 1749  
 Operator ID: Instrument ID: WC\_IonChrom7  
 Method: \\ChromNA\Denver\ChromData\WC\_IonChrom7\20180720-72165.b\Anions\_IC7.m  
 Limit Group: Wet - Anions  
 Last Update: 23-Jul-2018 09:24:33 Calib Date: 05-Jun-2018 12:02:00  
 Integrator: Falcon  
 Quant Method: External Standard Quant By: Initial Calibration  
 Last ICAL File: \\ChromNA\Denver\ChromData\WC\_IonChrom7\20180605-70676.b\07.0000.d  
 Column 1 : Det: 0005  
 Process Host: CTX0306

First Level Reviewer: jewellc Date: 20-Jul-2018 13:46:34

Compound	RT (min.)	Exp RT (min.)	Dlt RT (min.)	Response	Cal Amt ug/ml	OnCol Amt ug/ml	Flags
1 Fluoride	2.008	2.000	0.008	144408735	5.00	5.21	
2 Chloride	3.167	3.158	0.009	1664620997	100.0	97.8	
3 Nitrite as N	3.667	3.658	0.009	214431473	5.00	5.12	
4 Bromide	5.750	5.742	0.008	37207262	5.00	4.94	
5 Nitrate as N	6.192	6.183	0.009	220923785	5.00	4.98	
6 Sulfate	8.933	8.933	0.000	1316470664	100.0	103.8	a
7 Orthophosphate as P	11.817	11.775	0.042	70059071	5.00	4.19	

### QC Flag Legend

Review Flags

a - User Assigned ID

### Reagents:

IC LCS\_01288

Amount Added: 5.00

Units: mL


Report Date: 23-Jul-2018 09:24:52

Chrom Revision: 2.3 19-Jul-2018 15:14:50

## TestAmerica Denver

Data File: \\ChromNA\\Denver\\ChromData\\WC\_IonChrom7\\20180720-72165.b\\11.0000.d  
Injection Date: 20-Jul-2018 13:12:00 Instrument ID: WC\_IonChrom7  
Lims ID: lcs Operator ID:  
Client ID:  
Injection Vol: 25.0 ul ALS Bottle#: 0  
Method: Anions\_IC7 Dil. Factor: 1.0000  
Limit Group: Wet - Anions

Worklist Smp#: 11



TestAmerica Denver  
Target Compound Quantitation Report

Data File: \\ChromNA\Denver\ChromData\WC\_IonChrom7\20180720-72165.b\12.0000.d  
 Lims ID: lcsd  
 Client ID:  
 Sample Type: LCSD  
 Inject. Date: 20-Jul-2018 13:30:00 ALS Bottle#: 0 Worklist Smp#: 12  
 Injection Vol: 25.0 ul Dil. Factor: 1.0000  
 Sample Info: 280-0072165-012  
 Misc. Info.: 12  
 Operator ID: Instrument ID: WC\_IonChrom7  
 Method: \\ChromNA\Denver\ChromData\WC\_IonChrom7\20180720-72165.b\Anions\_IC7.m  
 Limit Group: Wet - Anions  
 Last Update: 23-Jul-2018 09:24:33 Calib Date: 05-Jun-2018 12:02:00  
 Integrator: Falcon  
 Quant Method: External Standard Quant By: Initial Calibration  
 Last ICAL File: \\ChromNA\Denver\ChromData\WC\_IonChrom7\20180605-70676.b\07.0000.d  
 Column 1 : Det: 0005  
 Process Host: CTX0306

First Level Reviewer: jewellc Date: 20-Jul-2018 14:54:44

Compound	RT (min.)	Exp RT (min.)	Dlt RT (min.)	Response	Cal Amt ug/ml	OnCol Amt ug/ml	Flags
1 Fluoride	2.000	2.000	0.000	143945092	5.00	5.19	
2 Chloride	3.167	3.158	0.009	1664030202	100.0	97.8	
3 Nitrite as N	3.658	3.658	0.000	214155613	5.00	5.12	
4 Bromide	5.742	5.742	0.000	37152893	5.00	4.94	
5 Nitrate as N	6.183	6.183	0.000	220670642	5.00	4.98	
6 Sulfate	8.933	8.933	0.000	1308255599	100.0	103.2	a
7 Orthophosphate as P	11.783	11.775	0.008	76147924	5.00	4.52	

### QC Flag Legend

Review Flags

a - User Assigned ID

### Reagents:

IC LCS\_01288

Amount Added: 5.00

Units: mL


Report Date: 23-Jul-2018 09:24:54

Chrom Revision: 2.3 19-Jul-2018 15:14:50

## TestAmerica Denver

Data File: \\ChromNA\\Denver\\ChromData\\WC\_IonChrom7\\20180720-72165.b\\12.0000.d  
Injection Date: 20-Jul-2018 13:30:00 Instrument ID: WC\_IonChrom7  
Lims ID: lcSD Operator ID:  
Client ID:  
Injection Vol: 25.0 ul ALS Bottle#: 0  
Method: Anions\_IC7 Dil. Factor: 1.0000  
Limit Group: Wet - Anions

Worklist Smp#: 12



TestAmerica Denver  
Target Compound Quantitation Report

Data File: \\ChromNA\Denver\ChromData\WC\_IonChrom7\20180720-72165.b\13.0000.d  
 Lims ID: mb  
 Client ID:  
 Sample Type: MB  
 Inject. Date: 20-Jul-2018 13:47:00 ALS Bottle#: 0 Worklist Smp#: 13  
 Injection Vol: 25.0 ul Dil. Factor: 1.0000  
 Sample Info: 280-0072165-013  
 Misc. Info.: 13 1493  
 Operator ID: Instrument ID: WC\_IonChrom7  
 Method: \\ChromNA\Denver\ChromData\WC\_IonChrom7\20180720-72165.b\Anions\_IC7.m  
 Limit Group: Wet - Anions  
 Last Update: 23-Jul-2018 09:24:33 Calib Date: 05-Jun-2018 12:02:00  
 Integrator: Falcon  
 Quant Method: External Standard Quant By: Initial Calibration  
 Last ICal File: \\ChromNA\Denver\ChromData\WC\_IonChrom7\20180605-70676.b\07.0000.d  
 Column 1 : Det: 0005  
 Process Host: CTX0306

Compound	RT (min.)	Exp RT (min.)	Dlt RT (min.)	Response	Cal Amt ug/ml	OnCol Amt ug/ml	Flags
1 Fluoride	1.983	2.000	-0.017	313211		-0.0283	
2 Chloride	3.167	3.158	0.009	310434		-0.0108	
3 Nitrite as N		3.658				ND	
4 Bromide		5.742				ND	
5 Nitrate as N	6.350	6.183	0.167	111962		0.008274	
6 Sulfate	8.333	8.933	-0.600	3431132		0.0899	
7 Orthophosphate as P		11.775				ND	


Report Date: 23-Jul-2018 09:24:56

Chrom Revision: 2.3 19-Jul-2018 15:14:50

## TestAmerica Denver

Data File: \\ChromNA\\Denver\\ChromData\\WC\_IonChrom7\\20180720-72165.b\\13.0000.d  
Injection Date: 20-Jul-2018 13:47:00 Instrument ID: WC\_IonChrom7  
Lims ID: mb Operator ID:  
Client ID:  
Injection Vol: 25.0 ul ALS Bottle#: 0  
Method: Anions\_IC7 Dil. Factor: 1.0000  
Limit Group: Wet - Anions

Worklist Smp#: 13



TestAmerica Denver  
Target Compound Quantitation Report

Data File: \\ChromNA\Denver\ChromData\WC\_IonChrom7\20180720-72165.b\14.0000.d  
 Lims ID: ccv  
 Client ID:  
 Sample Type: CCV  
 Inject. Date: 20-Jul-2018 14:05:00 ALS Bottle#: 0 Worklist Smp#: 14  
 Injection Vol: 25.0 ul Dil. Factor: 1.0000  
 Sample Info: 280-0072165-014  
 Misc. Info.: 14  
 Operator ID: Instrument ID: WC\_IonChrom7  
 Sublist: chrom-Anions\_IC7\*sub1  
 Method: \\ChromNA\Denver\ChromData\WC\_IonChrom7\20180720-72165.b\Anions\_IC7.m  
 Limit Group: Wet - Anions  
 Last Update: 23-Jul-2018 09:24:57 Calib Date: 05-Jun-2018 12:02:00  
 Integrator: Falcon  
 Quant Method: External Standard Quant By: Initial Calibration  
 Last ICal File: \\ChromNA\Denver\ChromData\WC\_IonChrom7\20180605-70676.b\07.0000.d  
 Column 1 : Det: 0005  
 Process Host: CTX0306

First Level Reviewer: jewelc Date: 20-Jul-2018 14:55:23

Compound	RT (min.)	Exp RT (min.)	Dlt RT (min.)	Response	Cal Amt ug/ml	OnCol Amt ug/ml	Flags
1 Fluoride	2.000	2.000	0.000	143472992	5.00	5.17	
2 Chloride	3.158	3.158	0.000	1661324016	100.0	97.6	
3 Nitrite as N	3.658	3.658	0.000	214425652	5.00	5.12	
4 Bromide	5.742	5.742	0.000	37092728	5.00	4.93	
5 Nitrate as N	6.183	6.183	0.000	220452491	5.00	4.97	
6 Sulfate	8.933	8.933	0.000	1306135844	100.0	103.0	a
7 Orthophosphate as P	11.775	11.775	0.000	76049633	5.00	4.52	

### QC Flag Legend

Review Flags

a - User Assigned ID

### Reagents:

IC LCS\_01288

Amount Added: 5.00

Units: mL

Report Date: 23-Jul-2018 09:24:57

Chrom Revision: 2.3 19-Jul-2018 15:14:50


TestAmerica Denver

Data File: \\ChromNA\\Denver\\ChromData\\WC\_IonChrom7\\20180720-72165.b\\14.0000.d  
Injection Date: 20-Jul-2018 14:05:00 Instrument ID: WC\_IonChrom7  
Lims ID: ccv Operator ID:  
Client ID:  
Injection Vol: 25.0 ul Worklist Smp#: 14  
Method: Anions\_IC7 Dil. Factor: 1.0000  
Limit Group: Wet - Anions

Operator ID:  
Worklist Smp#: 14

ALS Bottle#: 0

14.0000, 14.0000.d



TestAmerica Denver  
Target Compound Quantitation Report

Data File: \\ChromNA\Denver\ChromData\WC\_IonChrom7\20180720-72165.b\15.0000.d  
 Lims ID: ccb  
 Client ID:  
 Sample Type: CCB  
 Inject. Date: 20-Jul-2018 14:23:00 ALS Bottle#: 0 Worklist Smp#: 15  
 Injection Vol: 25.0 ul Dil. Factor: 1.0000  
 Sample Info: 280-0072165-015  
 Misc. Info.: 15 2424  
 Operator ID: Instrument ID: WC\_IonChrom7  
 Method: \\ChromNA\Denver\ChromData\WC\_IonChrom7\20180720-72165.b\Anions\_IC7.m  
 Limit Group: Wet - Anions  
 Last Update: 23-Jul-2018 09:24:57 Calib Date: 05-Jun-2018 12:02:00  
 Integrator: Falcon  
 Quant Method: External Standard Quant By: Initial Calibration  
 Last ICal File: \\ChromNA\Denver\ChromData\WC\_IonChrom7\20180605-70676.b\07.0000.d  
 Column 1 : Det: 0005  
 Process Host: CTX0306

Compound	RT (min.)	Exp RT (min.)	Dlt RT (min.)	Response	Cal Amt ug/ml	OnCol Amt ug/ml	Flags
1 Fluoride	1.983	2.000	-0.017	320911		-0.0280	
2 Chloride	3.175	3.158	0.017	242344		-0.0148	
3 Nitrite as N		3.658				ND	
4 Bromide		5.742				ND	
5 Nitrate as N		6.183				ND	
6 Sulfate	8.325	8.933	-0.608	3091479		0.0631	
7 Orthophosphate as P		11.775				ND	


Report Date: 23-Jul-2018 09:24:59

Chrom Revision: 2.3 19-Jul-2018 15:14:50

TestAmerica Denver

Data File: \\ChromNA\\Denver\\ChromData\\WC\_IonChrom7\\20180720-72165.b\\15.0000.d  
Injection Date: 20-Jul-2018 14:23:00 Instrument ID: WC\_IonChrom7  
Lims ID: ccb Operator ID:  
Client ID:  
Injection Vol: 25.0 ul ALS Bottle#: 0  
Method: Anions\_IC7 Dil. Factor: 1.0000  
Limit Group: Wet - Anions

Worklist Smp#: 15



TestAmerica Denver  
Target Compound Quantitation Report

Data File: \\ChromNA\Denver\ChromData\WC\_IonChrom7\20180720-72165.b\16.0000.d  
 Lims ID: mrl  
 Client ID:  
 Sample Type: MRL  
 Inject. Date: 20-Jul-2018 17:38:00 ALS Bottle#: 0 Worklist Smp#: 16  
 Injection Vol: 25.0 ul Dil. Factor: 1.0000  
 Sample Info: 280-0072165-016  
 Misc. Info.: 30076 2424  
 Operator ID: Instrument ID: WC\_IonChrom7  
 Method: \\ChromNA\Denver\ChromData\WC\_IonChrom7\20180720-72165.b\Anions\_IC7.m  
 Limit Group: Wet - Anions  
 Last Update: 23-Jul-2018 10:54:29 Calib Date: 05-Jun-2018 12:02:00  
 Integrator: Falcon  
 Quant Method: External Standard Quant By: Initial Calibration  
 Last ICal File: \\ChromNA\Denver\ChromData\WC\_IonChrom7\20180605-70676.b\07.0000.d  
 Column 1 : Det: 0005  
 Process Host: CTX0306

Compound	RT (min.)	Exp RT (min.)	Dlt RT (min.)	Response	Cal Amt ug/ml	OnCol Amt ug/ml	Flags
1 Fluoride	1.992	2.000	-0.008	6954375	0.2000	0.2130	
2 Chloride	3.183	3.158	0.025	46549247	2.50	2.71	
3 Nitrite as N	3.667	3.658	0.009	9775441	0.2000	0.2227	
4 Bromide	5.800	5.742	0.058	1364643	0.2000	0.2208	
5 Nitrate as N	6.358	6.183	0.175	9624887	0.2000	0.2226	
6 Sulfate	9.225	8.933	0.292	45380067	2.50	3.40	
7 Orthophosphate as P		11.775			ND	ND	

### QC Flag Legend

#### Processing Flags

ND - Not Detected or Marked ND

#### Reagents:

IC CAL cl/so4_00208	Amount Added: 0.05	Units: mL
IC Cal low_00385	Amount Added: 0.02	Units: mL


Report Date: 23-Jul-2018 10:54:36

Chrom Revision: 2.3 19-Jul-2018 15:14:50

TestAmerica Denver

Data File: \\ChromNA\\Denver\\ChromData\\WC\_IonChrom7\\20180720-72165.b\\16.0000.d  
Injection Date: 20-Jul-2018 17:38:00 Instrument ID: WC\_IonChrom7  
Lims ID: mrl Operator ID:  
Client ID:  
Injection Vol: 25.0 ul ALS Bottle#: 0  
Method: Anions\_IC7 Dil. Factor: 1.0000  
Limit Group: Wet - Anions

Worklist Smp#: 16



TestAmerica Denver  
Target Compound Quantitation Report

Data File: \\ChromNA\Denver\ChromData\WC\_IonChrom7\20180720-72165.b\38.0000.d  
 Lims ID: ccv  
 Client ID:  
 Sample Type: CCV  
 Inject. Date: 21-Jul-2018 00:10:00 ALS Bottle#: 0 Worklist Smp#: 38  
 Injection Vol: 25.0 ul Dil. Factor: 1.0000  
 Sample Info: 280-0072165-038  
 Misc. Info.: 20773  
 Operator ID: Instrument ID: WC\_IonChrom7  
 Sublist: chrom-Anions\_IC7\*sub1  
 Method: \\ChromNA\Denver\ChromData\WC\_IonChrom7\20180720-72165.b\Anions\_IC7.m  
 Limit Group: Wet - Anions  
 Last Update: 23-Jul-2018 09:25:28 Calib Date: 05-Jun-2018 12:02:00  
 Integrator: Falcon  
 Quant Method: External Standard Quant By: Initial Calibration  
 Last ICal File: \\ChromNA\Denver\ChromData\WC\_IonChrom7\20180605-70676.b\07.0000.d  
 Column 1 : Det: 0005  
 Process Host: CTX0306

Compound	RT (min.)	Exp RT (min.)	Dlt RT (min.)	Response	Cal Amt ug/ml	OnCol Amt ug/ml	Flags
1 Fluoride	2.008	2.000	0.008	145939572	5.00	5.26	
2 Chloride	3.167	3.158	0.009	1683388181	100.0	98.9	
3 Nitrite as N	3.667	3.658	0.009	215078206	5.00	5.14	
4 Bromide	5.750	5.742	0.008	37473421	5.00	4.98	
5 Nitrate as N	6.200	6.183	0.017	227090726	5.00	5.12	
6 Sulfate	8.925	8.933	-0.008	1341423430	100.0	105.8	
7 Orthophosphate as P	11.900	11.775	0.125	56824689	5.00	3.46	

**Reagents:**

IC LCS\_01288

Amount Added: 5.00

Units: mL


Report Date: 23-Jul-2018 09:25:29

Chrom Revision: 2.3 19-Jul-2018 15:14:50

## TestAmerica Denver

Data File: \\ChromNA\\Denver\\ChromData\\WC\_IonChrom7\\20180720-72165.b\\38.0000.d  
Injection Date: 21-Jul-2018 00:10:00 Instrument ID: WC\_IonChrom7  
Lims ID: ccv Operator ID:  
Client ID:  
Injection Vol: 25.0 ul Worklist Smp#: 38  
Method: Anions\_IC7 Dil. Factor: 1.0000  
Limit Group: Wet - Anions

38.0000, 38.0000.d



TestAmerica Denver  
Target Compound Quantitation Report

Data File: \\ChromNA\Denver\ChromData\WC\_IonChrom7\20180720-72165.b\39.0000.d  
 Lims ID: ccb  
 Client ID:  
 Sample Type: CCB  
 Inject. Date: 21-Jul-2018 00:28:00 ALS Bottle#: 0 Worklist Smp#: 39  
 Injection Vol: 25.0 ul Dil. Factor: 1.0000  
 Sample Info: 280-0072165-039  
 Misc. Info.: 29411  
 Operator ID: Instrument ID: WC\_IonChrom7  
 Method: \\ChromNA\Denver\ChromData\WC\_IonChrom7\20180720-72165.b\Anions\_IC7.m  
 Limit Group: Wet - Anions  
 Last Update: 23-Jul-2018 09:25:28 Calib Date: 05-Jun-2018 12:02:00  
 Integrator: Falcon  
 Quant Method: External Standard Quant By: Initial Calibration  
 Last ICal File: \\ChromNA\Denver\ChromData\WC\_IonChrom7\20180605-70676.b\07.0000.d  
 Column 1 : Det: 0005  
 Process Host: CTX0306

Compound	RT (min.)	Exp RT (min.)	Dlt RT (min.)	Response	Cal Amt ug/ml	OnCol Amt ug/ml	Flags
1 Fluoride	1.975	2.000	-0.025	202964	-0.0323		
2 Chloride	3.167	3.158	0.009	209493	-0.0167		
3 Nitrite as N		3.658			ND		
4 Bromide		5.742			ND		
5 Nitrate as N		6.183			ND		
6 Sulfate	9.258	8.933	0.325	1487378	-0.0636		
7 Orthophosphate as P		11.775			ND		


Report Date: 23-Jul-2018 09:25:30

Chrom Revision: 2.3 19-Jul-2018 15:14:50

## TestAmerica Denver

Data File: \\ChromNA\\Denver\\ChromData\\WC\_IonChrom7\\20180720-72165.b\\39.0000.d  
Injection Date: 21-Jul-2018 00:28:00 Instrument ID: WC\_IonChrom7  
Lims ID: ccb Operator ID:  
Client ID:  
Injection Vol: 25.0 ul ALS Bottle#: 0  
Method: Anions\_IC7 Dil. Factor: 1.0000  
Limit Group: Wet - Anions

Worklist Smp#: 39



TestAmerica Denver  
Target Compound Quantitation Report

Data File: \\ChromNA\Denver\ChromData\WC\_IonChrom7\20180720-72165.b\40.0000.d  
 Lims ID: 280-111519-H-14  
 Client ID: LL1mw-083-062718-GW  
 Sample Type: Client  
 Inject. Date: 21-Jul-2018 00:46:00 ALS Bottle#: 0 Worklist Smp#: 40  
 Injection Vol: 25.0 ul Dil. Factor: 1.0000  
 Sample Info: 280-0072165-040  
 Misc. Info.: 23961 F  
 Operator ID: Instrument ID: WC\_IonChrom7  
 Method: \\ChromNA\Denver\ChromData\WC\_IonChrom7\20180720-72165.b\Anions\_IC7.m  
 Limit Group: Wet - Anions  
 Last Update: 03-Oct-2018 18:11:56 Calib Date: 05-Jun-2018 12:02:00  
 Integrator: Falcon  
 Quant Method: External Standard Quant By: Initial Calibration  
 Last ICAL File: \\ChromNA\Denver\ChromData\WC\_IonChrom7\20180605-70676.b\07.0000.d  
 Column 1 : Det: 0005  
 Process Host: CTX0317

Compound	RT (min.)	Exp RT (min.)	Dlt RT (min.)	Response	OnCol Amt ug/ml	Flags
1 Fluoride	1.992	2.000	-0.008	1081914	-0.000352	
2 Chloride	3.183	3.158	0.025	58620578	3.42	
3 Nitrite as N		3.658			ND	
4 Bromide	5.775	5.742	0.033	151028	0.0609	
5 Nitrate as N	6.342	6.183	0.159	13503212	0.3100	
6 Sulfate	8.842	8.933	-0.091	1866038202	147.2	
7 Orthophosphate as P		11.775			ND	

Report Date: 03-Oct-2018 18:12:05


Chrom Revision: 2.3 19-Jul-2018 15:14:50

TestAmerica Denver

Data File: \\ChromNA\\Denver\\ChromData\\WC\_IonChrom7\\20180720-72165.b\\40.0000.d  
Injection Date: 21-Jul-2018 00:46:00 Instrument ID: WC\_IonChrom7  
Lims ID: 280-111519-H-14 Lab Sample ID: 280-111519-14  
Client ID: LL1mw-083-062718-GW  
Injection Vol: 25.0 ul Dil. Factor: 1.0000  
Method: Anions\_IC7 Limit Group: Wet - Anions

Operator ID:  
Worklist Smp#: 40

ALS Bottle#: 0



TestAmerica Denver  
Target Compound Quantitation Report

Data File: \\ChromNA\Denver\ChromData\WC\_IonChrom7\20180720-72165.b\41.0000.d  
 Lims ID: 280-111519-D-16  
 Client ID: RQLmw-011-062818-GW  
 Sample Type: Client  
 Inject. Date: 21-Jul-2018 01:04:00 ALS Bottle#: 0 Worklist Smp#: 41  
 Injection Vol: 25.0 ul Dil. Factor: 1.0000  
 Sample Info: 280-0072165-041  
 Misc. Info.: 12052 F  
 Operator ID: Instrument ID: WC\_IonChrom7  
 Method: \\ChromNA\Denver\ChromData\WC\_IonChrom7\20180720-72165.b\Anions\_IC7.m  
 Limit Group: Wet - Anions  
 Last Update: 03-Oct-2018 18:11:56 Calib Date: 05-Jun-2018 12:02:00  
 Integrator: Falcon  
 Quant Method: External Standard Quant By: Initial Calibration  
 Last ICal File: \\ChromNA\Denver\ChromData\WC\_IonChrom7\20180605-70676.b\07.0000.d  
 Column 1 : Det: 0005  
 Process Host: CTX0317

Compound	RT (min.)	Exp RT (min.)	Dlt RT (min.)	Response	OnCol Amt ug/ml	Flags
1 Fluoride	1.992	2.000	-0.008	1478552	0.0141	
2 Chloride	3.175	3.158	0.017	19857984	1.14	
3 Nitrite as N		3.658			ND	
4 Bromide		5.742			ND	
5 Nitrate as N	6.317	6.183	0.134	165107	0.009472	
6 Sulfate	8.817	8.933	-0.116	2031116575	160.3	
7 Orthophosphate as P		11.775			ND	

Report Date: 03-Oct-2018 18:12:06


Chrom Revision: 2.3 19-Jul-2018 15:14:50

TestAmerica Denver

Data File: \\ChromNA\\Denver\\ChromData\\WC\_IonChrom7\\20180720-72165.b\\41.0000.d  
Injection Date: 21-Jul-2018 01:04:00 Instrument ID: WC\_IonChrom7  
Lims ID: 280-111519-D-16 Lab Sample ID: 280-111519-16  
Client ID: RQLmw-011-062818-GW  
Injection Vol: 25.0 ul Dil. Factor: 1.0000  
Method: Anions\_IC7 Limit Group: Wet - Anions

Operator ID:  
Worklist Smp#: 41

ALS Bottle#: 0



TestAmerica Denver  
Target Compound Quantitation Report

Data File: \\ChromNA\Denver\ChromData\WC\_IonChrom7\20180720-72165.b\50.0000.d  
 Lims ID: ccv  
 Client ID:  
 Sample Type: CCV  
 Inject. Date: 21-Jul-2018 03:44:00 ALS Bottle#: 0 Worklist Smp#: 50  
 Injection Vol: 25.0 ul Dil. Factor: 1.0000  
 Sample Info: 280-0072165-050  
 Misc. Info.: 7220  
 Operator ID: Instrument ID: WC\_IonChrom7  
 Sublist: chrom-Anions\_IC7\*sub1  
 Method: \\ChromNA\Denver\ChromData\WC\_IonChrom7\20180720-72165.b\Anions\_IC7.m  
 Limit Group: Wet - Anions  
 Last Update: 23-Jul-2018 09:25:42 Calib Date: 05-Jun-2018 12:02:00  
 Integrator: Falcon  
 Quant Method: External Standard Quant By: Initial Calibration  
 Last ICal File: \\ChromNA\Denver\ChromData\WC\_IonChrom7\20180605-70676.b\07.0000.d  
 Column 1 : Det: 0005  
 Process Host: CTX0306

Compound	RT (min.)	Exp RT (min.)	Dlt RT (min.)	Response	Cal Amt ug/ml	OnCol Amt ug/ml	Flags
1 Fluoride	2.000	2.000	0.000	149078512	5.00	5.38	
2 Chloride	3.167	3.158	0.009	1714295503	100.0	100.7	
3 Nitrite as N	3.658	3.658	0.000	215407099	5.00	5.15	
4 Bromide	5.758	5.742	0.016	37973082	5.00	5.04	
5 Nitrate as N	6.208	6.183	0.025	225132281	5.00	5.08	
6 Sulfate	8.892	8.933	-0.041	1320925633	100.0	104.2	
7 Orthophosphate as P	11.608	11.775	-0.167	99952798	5.00	5.84	

**Reagents:**

IC LCS\_01288 Amount Added: 5.00 Units: mL


Report Date: 23-Jul-2018 09:25:43

Chrom Revision: 2.3 19-Jul-2018 15:14:50

TestAmerica Denver

Data File: \\ChromNA\\Denver\\ChromData\\WC\_IonChrom7\\20180720-72165.b\\50.0000.d  
Injection Date: 21-Jul-2018 03:44:00 Instrument ID: WC\_IonChrom7  
Lims ID: ccv Operator ID:  
Client ID:  
Injection Vol: 25.0 ul ALS Bottle#: 0  
Method: Anions\_IC7 Dil. Factor: 1.0000  
Limit Group: Wet - Anions

Worklist Smp#: 50



TestAmerica Denver  
Target Compound Quantitation Report

Data File: \\ChromNA\Denver\ChromData\WC\_IonChrom7\20180720-72165.b\51.0000.d  
 Lims ID: ccb  
 Client ID:  
 Sample Type: CCB  
 Inject. Date: 21-Jul-2018 04:02:00 ALS Bottle#: 0 Worklist Smp#: 51  
 Injection Vol: 25.0 ul Dil. Factor: 1.0000  
 Sample Info: 280-0072165-051  
 Misc. Info.: 9752  
 Operator ID: Instrument ID: WC\_IonChrom7  
 Method: \\ChromNA\Denver\ChromData\WC\_IonChrom7\20180720-72165.b\Anions\_IC7.m  
 Limit Group: Wet - Anions  
 Last Update: 23-Jul-2018 09:25:42 Calib Date: 05-Jun-2018 12:02:00  
 Integrator: Falcon  
 Quant Method: External Standard Quant By: Initial Calibration  
 Last ICal File: \\ChromNA\Denver\ChromData\WC\_IonChrom7\20180605-70676.b\07.0000.d  
 Column 1 : Det: 0005  
 Process Host: CTX0306

Compound	RT (min.)	Exp RT (min.)	Dlt RT (min.)	Response	Cal Amt ug/ml	OnCol Amt ug/ml	Flags
1 Fluoride	1.983	2.000	-0.017	497649	-0.0216		
2 Chloride	3.167	3.158	0.009	161758	-0.0195		
3 Nitrite as N		3.658			ND		
4 Bromide		5.742			ND		
5 Nitrate as N		6.183			ND		
6 Sulfate	9.217	8.933	0.284	1006339	-0.1016		
7 Orthophosphate as P		11.775			ND		


Report Date: 23-Jul-2018 09:25:44

Chrom Revision: 2.3 19-Jul-2018 15:14:50

TestAmerica Denver

Data File: \\ChromNA\\Denver\\ChromData\\WC\_IonChrom7\\20180720-72165.b\\51.0000.d  
Injection Date: 21-Jul-2018 04:02:00 Instrument ID: WC\_IonChrom7  
Lims ID: ccb Operator ID:  
Client ID:  
Injection Vol: 25.0 ul ALS Bottle#: 0  
Method: Anions\_IC7 Dil. Factor: 1.0000  
Limit Group: Wet - Anions

Worklist Smp#: 51



# **Shipping and Receiving Documents**



Chain of Custody Record  
TestAmerica Denver  
19955 Yarrow Street

4955 Yarrow Street  
Arvada CO 80002

Arvada, CO 80002  
Phone (303) 736-0100 Fax (303) 431-7171

TestAmerica

THE CHASE IN CANTONMENT TERRITORIES



**TestAmerica Denver**

4955 Yarrow Street  
Arvada, CO 80002  
Phone (303) 736-0100 Fax (303) 431-7711

**Chain of Custody Record**

**TestAmerica**  
THE LEADING TEST LABORATORY

\*Hazardous Substance Test

Client Contact: Danyelle Phillips Company: Cardno TEC, Inc.

Address: 1658 Cole Boulevard Suite 190 City: Golden State, Zip: CO 80401 Phone: 434-906-2085 Email: Danyelle.Phillips@cardnotecs.com Project Name: Ravenna, OH Site: Ravenna

Job #:

Sample #: 434-906-2085 Lab F/T: McEntee, Patrick J E-Mail: patrick.mcintee@testamericainc.com

Carri# Tracking No(s):

Comments/MSD (Yes or No):

Field Filtered Sample (Yes or No):

Preservation Code: A - HCl B - NaOH C - Zn Acetate D - Nitric Acid E - NaISO4 F - MeOH G - Acetone H - Ascorbic Acid I - Ica J - DI Water K - EDTA L - ELIA Other:

Total Number of Contaminants:

Special Instructions/Note:

QC - dep

Analysis Requested:

Due Date Requested:

TAT Requested (days): 20 Business Days

PO #:

WO #: 076003.009.0111

Project #: 28014271

SSOW#:

Matrix (Waste, Ground, Unknown, Interfacing, Grab):

Sample Date:

Sample Time:

Sample Type (C=Comp, G=grab):

Preservation Code:

Time:

Method of Shipment:

Date/Time Received by:

Date/Time Disposal By Lab:

Archive For: months

Return To Client:

Special Instructions/QC Requirements:

Possible Hazard Identification:

Non-Hazard  Flammable  Skin Irritant  Poison B  Unknown  Radiological

Deliverable Requested: I, II, III, IV, Other (specify)

Empty Kit Relinquished by:

Relinquished by: Received by: Date/Time:

Relinquished by: Received by: Date/Time:

Relinquished by: Received by: Date/Time:

Custody Seals intact:  Yes  No

Cooler Temperature(s) °C and Other Remarks:

TestAmerica Denver  
4955 Yarrow Street  
Arvada, CO 80002  
Phone (303) 736-0100 Fax (303) 431-7771

## Chain of Custody Record

PRINTED IN INK OR COMPUTER PRINTED

TestAmerica

Client Information		Sampler:	Lab FM:	Analysis Requested																						
		Phone:	E-Mail:	Carrier Tracking No(s):																						
Cardno TEC, Inc		Jep	McEnlee, Patrick J																							
Address:																										
City:	1658 Colle Boulevard Suite 190			TAT Requested (days): <b>20 Business Days</b>																						
State, Zip:																										
CO 80401																										
Phone:																										
434-906-2085																										
Email:	Danyelle Phillips@cardno-qs.com																									
Project Name:	076003.009.011																									
Site:	Ravenna, OH -																									
															Special Instructions/Note: <i>QC - dep</i>											
															Batch Number of Container:											
															<b>MS/MSD</b>											
Sample Identification		Preservation Code:			Sample Date	Sample Time	Sample Type	Matrix	Custodian																	
RQLWW-011-062818-GW		6-28-18 1151 G W MN					(C=Comp, G=Grab)	(Waste, Soil, Drum, etc.)	N	N	N	N	N	N	N	N	N	N	N	N	N					
Possible Hazard Identification					<input type="checkbox"/> Non-Hazard	<input type="checkbox"/> Flammable	<input type="checkbox"/> Skin Irritant	<input type="checkbox"/> Poison B	<input type="checkbox"/> Unknown	<input type="checkbox"/> Radiological	Time:															
Deliverable Requested: I, II, III, IV, Other (specify)					Date/Time:																					
Empty Kit Relinquished by:					Received by:																					
Relinquished by:					Time:																					
Relinquished by:					Date/Time:																					
Custody Seals intact:		Custody Seal No:			Date/Time:																					
△ Yes		△ No			Cooler Temperature(s) °C and Other Remarks:																					

**TestAmerica Denver**  
4955 Yarrow Street  
Arvada, CO 80010 Fax (303) 431-7171

**Chain of Custody Record**

Client Information		Sampler: SC 93C	Lab FM: McEntee, Patrick J	Carrier Tracking No(s):
Client Contact:	Phone: 303-906-2085	E-Mail: patrick.mcnette@testamericainc.com		
Cardno TEC, Inc.				
Address:	1658 Cole Boulevard Suite 190			
City:		TAT Requested (days):	20 Business Days	
State, Zip:	CO 80401	PO #:		
Phone:	434-906-2085	WO #:	076003.009.011	
E-mail:	Danyelle Phillips@cardno-qs.com	Project #:	28014271	
Project Name:	Ravenna, OH - Site: Ravenna	SSCRN#:		
Field Filtered Sample (Yes or No)				
Preservation Code (Yes or No)				
Sample Identification	Sample Date	Sample Time	Sample Type (C=Const., G=Grab)	Matrix (Froster, O=Oil, B=Brine, S=Sediment, D=Drill Cuttings, N=N/A, A=Acetone, B/D=B/D, C/B/N=C/B/N, D=N/D)
LL1 mus - 084-0628148	0628180824	C	C	N
SC Fass - 004-0628148	0628180814	C	C	N
<i>drop</i>				
<b>Possible Hazard Identification</b> <input type="checkbox"/> Non-Hazard <input type="checkbox"/> Flammable <input type="checkbox"/> Skin Irritant <input type="checkbox"/> Poison B <input type="checkbox"/> Unknown <input type="checkbox"/> Radiological Deliverable Requested: I, II, III, IV, Other (specify)				
Empty Kit Reinquisition by: Received by: <i>John</i> Date/Time: <i>6-28-18</i> Company: <i>Cardno</i> Reinquisition by: <i>John</i> Date/Time: <i>6-28-18-1635</i> Company: <i>Cardno</i>				
Custody Seals intact: <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No Custody Seal No.: <i>220</i>				
Total Number of Contaminants: <i>5</i> Special Instructions/Note: <i>QC - dry</i>				
Preservation Codes: A - HCl   M - HBrine B - NaOH   N - None C - Zn Acetate   O - AnNaO2 D - Nitric Acid   P - Na2O4S E - NaHSO4   Q - Na2S2O3 F - MeOH   R - Na2S2O3 G - Ammonium   S - H2SO4 H - Ascorbic Acid   T - TSP Disodium Phosphate I - Ice   U - Acetone V - MCA W - pH 4.5 X - EDTA Y - EDA Z - other (specify)				
8010C-6020A/7470A - Total Metals 8010C-6020A/7470A - Dissolved Metals 8034 - Arsenic 8038 - Barium 8040 - Cadmium 8041C - Phenophenous 8042A - Arsenite and Sulfite 8043A - Nitrate (48 HOUR HOLD TIME) 8046B - Perchlorate 8054 - Silicate 8056A - Ammonium (Chloride and Sulfite) 8220B - Alkalinity 8270A - Hexavalent Chromium (24 HOUR HOLD TIME) 8270B - Sulfide 8270D - Arsenite 8270D - S-VOCs (LV) PAHs (LV) 8270D - SVOCs (LV) Sulfur 8270D - SVOCs (LV) 1,4-Dioxane 8270D - SVOCs (LV) 2 8270D - SVOCs (LV) 3 8270D - SVOCs (LV) 4 8270D - SVOCs (LV) 5 8270D - SVOCs (LV) 6 8270D - SVOCs (LV) 7 8270D - SVOCs (LV) 8 8270D - SVOCs (LV) 9 8270D - SVOCs (LV) 10 8270D - SVOCs (LV) 11 8270D - SVOCs (LV) 12 8270D - SVOCs (LV) 13 8270D - SVOCs (LV) 14 8270D - SVOCs (LV) 15 8270D - SVOCs (LV) 16 8270D - SVOCs (LV) 17 8270D - SVOCs (LV) 18 8270D - SVOCs (LV) 19 8270D - SVOCs (LV) 20 8270D - SVOCs (LV) 21 8270D - SVOCs (LV) 22 8270D - SVOCs (LV) 23 8270D - SVOCs (LV) 24 8270D - SVOCs (LV) 25 8270D - SVOCs (LV) 26 8270D - SVOCs (LV) 27 8270D - SVOCs (LV) 28 8270D - SVOCs (LV) 29 8270D - SVOCs (LV) 30 8270D - SVOCs (LV) 31 8270D - SVOCs (LV) 32 8270D - SVOCs (LV) 33 8270D - SVOCs (LV) 34 8270D - SVOCs (LV) 35 8270D - SVOCs (LV) 36 8270D - SVOCs (LV) 37 8270D - SVOCs (LV) 38 8270D - SVOCs (LV) 39 8270D - SVOCs (LV) 40 8270D - SVOCs (LV) 41 8270D - SVOCs (LV) 42 8270D - SVOCs (LV) 43 8270D - SVOCs (LV) 44 8270D - SVOCs (LV) 45 8270D - SVOCs (LV) 46 8270D - SVOCs (LV) 47 8270D - SVOCs (LV) 48 8270D - SVOCs (LV) 49 8270D - SVOCs (LV) 50 8270D - SVOCs (LV) 51 8270D - SVOCs (LV) 52 8270D - SVOCs (LV) 53 8270D - SVOCs (LV) 54 8270D - SVOCs (LV) 55 8270D - SVOCs (LV) 56 8270D - SVOCs (LV) 57 8270D - SVOCs (LV) 58 8270D - SVOCs (LV) 59 8270D - SVOCs (LV) 60 8270D - SVOCs (LV) 61 8270D - SVOCs (LV) 62 8270D - SVOCs (LV) 63 8270D - SVOCs (LV) 64 8270D - SVOCs (LV) 65 8270D - SVOCs (LV) 66 8270D - SVOCs (LV) 67 8270D - SVOCs (LV) 68 8270D - SVOCs (LV) 69 8270D - SVOCs (LV) 70 8270D - SVOCs (LV) 71 8270D - SVOCs (LV) 72 8270D - SVOCs (LV) 73 8270D - SVOCs (LV) 74 8270D - SVOCs (LV) 75 8270D - SVOCs (LV) 76 8270D - SVOCs (LV) 77 8270D - SVOCs (LV) 78 8270D - SVOCs (LV) 79 8270D - SVOCs (LV) 80 8270D - SVOCs (LV) 81 8270D - SVOCs (LV) 82 8270D - SVOCs (LV) 83 8270D - SVOCs (LV) 84 8270D - SVOCs (LV) 85 8270D - SVOCs (LV) 86 8270D - SVOCs (LV) 87 8270D - SVOCs (LV) 88 8270D - SVOCs (LV) 89 8270D - SVOCs (LV) 90 8270D - SVOCs (LV) 91 8270D - SVOCs (LV) 92 8270D - SVOCs (LV) 93 8270D - SVOCs (LV) 94 8270D - SVOCs (LV) 95 8270D - SVOCs (LV) 96 8270D - SVOCs (LV) 97 8270D - SVOCs (LV) 98 8270D - SVOCs (LV) 99 8270D - SVOCs (LV) 100 8270D - SVOCs (LV) 101 8270D - SVOCs (LV) 102 8270D - SVOCs (LV) 103 8270D - SVOCs (LV) 104 8270D - SVOCs (LV) 105 8270D - SVOCs (LV) 106 8270D - SVOCs (LV) 107 8270D - SVOCs (LV) 108 8270D - SVOCs (LV) 109 8270D - SVOCs (LV) 110 8270D - SVOCs (LV) 111 8270D - SVOCs (LV) 112 8270D - SVOCs (LV) 113 8270D - SVOCs (LV) 114 8270D - SVOCs (LV) 115 8270D - SVOCs (LV) 116 8270D - SVOCs (LV) 117 8270D - SVOCs (LV) 118 8270D - SVOCs (LV) 119 8270D - SVOCs (LV) 120 8270D - SVOCs (LV) 121 8270D - SVOCs (LV) 122 8270D - SVOCs (LV) 123 8270D - SVOCs (LV) 124 8270D - SVOCs (LV) 125 8270D - SVOCs (LV) 126 8270D - SVOCs (LV) 127 8270D - SVOCs (LV) 128 8270D - SVOCs (LV) 129 8270D - SVOCs (LV) 130 8270D - SVOCs (LV) 131 8270D - SVOCs (LV) 132 8270D - SVOCs (LV) 133 8270D - SVOCs (LV) 134 8270D - SVOCs (LV) 135 8270D - SVOCs (LV) 136 8270D - SVOCs (LV) 137 8270D - SVOCs (LV) 138 8270D - SVOCs (LV) 139 8270D - SVOCs (LV) 140 8270D - SVOCs (LV) 141 8270D - SVOCs (LV) 142 8270D - SVOCs (LV) 143 8270D - SVOCs (LV) 144 8270D - SVOCs (LV) 145 8270D - SVOCs (LV) 146 8270D - SVOCs (LV) 147 8270D - SVOCs (LV) 148 8270D - SVOCs (LV) 149 8270D - SVOCs (LV) 150 8270D - SVOCs (LV) 151 8270D - SVOCs (LV) 152 8270D - SVOCs (LV) 153 8270D - SVOCs (LV) 154 8270D - SVOCs (LV) 155 8270D - SVOCs (LV) 156 8270D - SVOCs (LV) 157 8270D - SVOCs (LV) 158 8270D - SVOCs (LV) 159 8270D - SVOCs (LV) 160 8270D - SVOCs (LV) 161 8270D - SVOCs (LV) 162 8270D - SVOCs (LV) 163 8270D - SVOCs (LV) 164 8270D - SVOCs (LV) 165 8270D - SVOCs (LV) 166 8270D - SVOCs (LV) 167 8270D - SVOCs (LV) 168 8270D - SVOCs (LV) 169 8270D - SVOCs (LV) 170 8270D - SVOCs (LV) 171 8270D - SVOCs (LV) 172 8270D - SVOCs (LV) 173 8270D - SVOCs (LV) 174 8270D - SVOCs (LV) 175 8270D - SVOCs (LV) 176 8270D - SVOCs (LV) 177 8270D - SVOCs (LV) 178 8270D - SVOCs (LV) 179 8270D - SVOCs (LV) 180 8270D - SVOCs (LV) 181 8270D - SVOCs (LV) 182 8270D - SVOCs (LV) 183 8270D - SVOCs (LV) 184 8270D - SVOCs (LV) 185 8270D - SVOCs (LV) 186 8270D - SVOCs (LV) 187 8270D - SVOCs (LV) 188 8270D - SVOCs (LV) 189 8270D - SVOCs (LV) 190 8270D - SVOCs (LV) 191 8270D - SVOCs (LV) 192 8270D - SVOCs (LV) 193 8270D - SVOCs (LV) 194 8270D - SVOCs (LV) 195 8270D - SVOCs (LV) 196 8270D - SVOCs (LV) 197 8270D - SVOCs (LV) 198 8270D - SVOCs (LV) 199 8270D - SVOCs (LV) 200 8270D - SVOCs (LV) 201 8270D - SVOCs (LV) 202 8270D - SVOCs (LV) 203 8270D - SVOCs (LV) 204 8270D - SVOCs (LV) 205 8270D - SVOCs (LV) 206 8270D - SVOCs (LV) 207 8270D - SVOCs (LV) 208 8270D - SVOCs (LV) 209 8270D - SVOCs (LV) 210 8270D - SVOCs (LV) 211 8270D - SVOCs (LV) 212 8270D - SVOCs (LV) 213 8270D - SVOCs (LV) 214 8270D - SVOCs (LV) 215 8270D - SVOCs (LV) 216 8270D - SVOCs (LV) 217 8270D - SVOCs (LV) 218 8270D - SVOCs (LV) 219 8270D - SVOCs (LV) 220 8270D - SVOCs (LV) 221 8270D - SVOCs (LV) 222 8270D - SVOCs (LV) 223 8270D - SVOCs (LV) 224 8270D - SVOCs (LV) 225 8270D - SVOCs (LV) 226 8270D - SVOCs (LV) 227 8270D - SVOCs (LV) 228 8270D - SVOCs (LV) 229 8270D - SVOCs (LV) 230 8270D - SVOCs (LV) 231 8270D - SVOCs (LV) 232 8270D - SVOCs (LV) 233 8270D - SVOCs (LV) 234 8270D - SVOCs (LV) 235 8270D - SVOCs (LV) 236 8270D - SVOCs (LV) 237 8270D - SVOCs (LV) 238 8270D - SVOCs (LV) 239 8270D - SVOCs (LV) 240 8270D - SVOCs (LV) 241 8270D - SVOCs (LV) 242 8270D - SVOCs (LV) 243 8270D - SVOCs (LV) 244 8270D - SVOCs (LV) 245 8270D - SVOCs (LV) 246 8270D - SVOCs (LV) 247 8270D - SVOCs (LV) 248 8270D - SVOCs (LV) 249 8270D - SVOCs (LV) 250 8270D - SVOCs (LV) 251 8270D - SVOCs (LV) 252 8270D - SVOCs (LV) 253 8270D - SVOCs (LV) 254 8270D - SVOCs (LV) 255 8270D - SVOCs (LV) 256 8270D - SVOCs (LV) 257 8270D - SVOCs (LV) 258 8270D - SVOCs (LV) 259 8270D - SVOCs (LV) 260 8270D - SVOCs (LV) 261 8270D - SVOCs (LV) 262 8270D - SVOCs (LV) 263 8270D - SVOCs (LV) 264 8270D - SVOCs (LV) 265 8270D - SVOCs (LV) 266 8270D - SVOCs (LV) 267 8270D - SVOCs (LV) 268 8270D - SVOCs (LV) 269 8270D - SVOCs (LV) 270 8270D - SVOCs (LV) 271 8270D - SVOCs (LV) 272 8270D - SVOCs (LV) 273 8270D - SVOCs (LV) 274 8270D - SVOCs (LV) 275 8270D - SVOCs (LV) 276 8270D - SVOCs (LV) 277 8270D - SVOCs (LV) 278 8270D - SVOCs (LV) 279 8270D - SVOCs (LV) 280 8270D - SVOCs (LV) 281 8270D - SVOCs (LV) 282 8270D - SVOCs (LV) 283 8270D - SVOCs (LV) 284 8270D - SVOCs (LV) 285 8270D - SVOCs (LV) 286 8270D - SVOCs (LV) 287 8270D - SVOCs (LV) 288 8270D - SVOCs (LV) 289 8270D - SVOCs (LV) 290 8270D - SVOCs (LV) 291 8270D - SVOCs (LV) 292 8270D - SVOCs (LV) 293 8270D - SVOCs (LV) 294 8270D - SVOCs (LV) 295 8270D - SVOCs (LV) 296 8270D - SVOCs (LV) 297 8270D - SVOCs (LV) 298 8270D - SVOCs (LV) 299 8270D - SVOCs (LV) 300 8270D - SVOCs (LV) 301 8270D - SVOCs (LV) 302 8270D - SVOCs (LV) 303 8270D - SVOCs (LV) 304 8270D - SVOCs (LV) 305 8270D - SVOCs (LV) 306 8270D - SVOCs (LV) 307 8270D - SVOCs (LV) 308 8270D - SVOCs (LV) 309 8270D - SVOCs (LV) 310 8270D - SVOCs (LV) 311 8270D - SVOCs (LV) 312 8270D - SVOCs (LV) 313 8270D - SVOCs (LV) 314 8270D - SVOCs (LV) 315 8270D - SVOCs (LV) 316 8270D - SVOCs (LV) 317 8270D - SVOCs (LV) 318 8270D - SVOCs (LV) 319 8270D - SVOCs (LV) 320 8270D - SVOCs (LV) 321 8270D - SVOCs (LV) 322 8270D - SVOCs (LV) 323 8270D - SVOCs (LV) 324 8270D - SVOCs (LV) 325 8270D - SVOCs (LV) 326 8270D - SVOCs (LV) 327 8270D - SVOCs (LV) 328 8270D - SVOCs (LV) 329 8270D - SVOCs (LV) 330 8270D - SVOCs (LV) 331 8270D - SVOCs (LV) 332 8270D - SVOCs (LV) 333 8270D - SVOCs (LV) 334 8270D - SVOCs (LV) 335 8270D - SVOCs (LV) 336 8270D - SVOCs (LV) 337 8270D - SVOCs (LV) 338 8270D - SVOCs (LV) 339 8270D - SVOCs (LV) 340 8270D - SVOCs (LV) 341 8270D - SVOCs (LV) 342 8270D - SVOCs (LV) 343 8270D - SVOCs (LV) 344 8270D - SVOCs (LV) 345 8270D - SVOCs (LV) 346 8270D - SVOCs (LV) 347 8270D - SVOCs (LV) 348 8270D - SVOCs (LV) 349 8270D - SVOCs (LV) 350 8270D - SVOCs (LV) 351 8270D - SVOCs (LV) 352 8270D - SVOCs (LV) 353 8270D - SVOCs (LV) 354 8270D - SVOCs (LV) 355 8270D - SVOCs (LV) 356 8270D - SVOCs (LV) 357 8270D - SVOCs (LV) 358 8270D - SVOCs (LV) 359 8270D - SVOCs (LV) 360 8270D - SVOCs (LV) 361 8270D - SVOCs (LV) 362 8270D - SVOCs (LV) 363 8270D - SVOCs (LV) 364 8270D - SVOCs (LV) 365 8270D - SVOCs (LV) 366 8270D - SVOCs (LV) 367 8270D - SVOCs (LV) 368 8270D - SVOCs (LV) 369 8270D - SVOCs (LV) 370 8270D - SVOCs (LV) 371 8270D - SVOCs (LV) 372 8270D - SVOCs (LV) 373 8270D - SVOCs (LV) 374 8270D - SVOCs (LV) 375 8270D - SVOCs (LV) 376 8270D - SVOCs (LV) 377 8270D - SVOCs (LV) 378 8270D - SVOCs (LV) 379 8270D - SVOCs (LV) 380 8270D - SVOCs (LV) 381 8270D - SVOCs (LV) 382 8270D - SVOCs (LV) 383 8270D - SVOCs (LV) 384 8270D - SVOCs (LV) 385 8270D - SVOCs (LV) 386 8270D - SVOCs (LV) 387 8270D - SVOCs (LV) 388 8270D - SVOCs (LV) 389 8270D - SVOCs (LV) 390 8270D - SVOCs (LV) 391 8270D - SVOCs (LV) 392 8270D - SVOCs (LV) 393 8270D - SVOCs (LV) 394 8270D - SVOCs (LV) 395 8270D - SVOCs (LV) 396 8270D - SVOCs (LV) 397 8270D - SVOCs (LV) 398 8270D - SVOCs (LV) 399 8270D - SVOCs (LV) 400				

**TestAmerica Denver**

4955 Yarrow Street  
Arvada, CO 80002  
Phone (303) 736-0100 Fax (303) 431-7171

**Chain of Custody Record**

TEST AMERICA INC. CUSTODY RECORD FORM 10-1992

Client Information		Sampler:	Lab P.M.	Carrier Tracking No(s):			
Client Contact:	Danyelle Phillips	Phone:	934-900-2085	McEntee, Patrick J	E-Mail:	patrick.mcintee@testamericadating.com	
Company:	Cardno TEC, Inc.	PO #:		Job #:			
Address:	1658 Cole Boulevard Suite 190	Due Date Requested:		Analysis Requested			
City:		TAT Requested (days):	20 Business Days				
State/Zip:							
CO 80401							
Phone:							
Email:							
Danyelle Phillips@cardno-acs.com							
Project Name:	076003.009.011	WO #:					
Ravenna, OH - Ravenna	28014271	Project #:					
Site SSOW#:		SSOW#:					
Preservation Codes:							
Sample Identification	Sample Date	Sample Time	Sample Type (C=comp, G=grab)	Matrix (Waste, Swallow, Ingestion, Animal)	Preservation Code:		
LL12mll-242-1-D-061715-C-062214	1140	C	C	W	8270D-SVOCs List 4		
LL12mll-242-062218-C-062214	1140	C	G	W	8270D-SVOCs List 2		
					8270D-SVOCs List 3		
					8270D-SVOCs List 1		
					8280B-VOCs		
					8292A-PCBs		
					8301B-Cyanide		
					8320B-Aromatic		
					7192A-Hexavalent Chromium (6 HOUR HOLD TIME)		
					2320B-Alkalinity		
					9056A-Anions (Chloride and Sulfate)		
					9334-Sulfide		
					9058A-Nitrile (8 HOUR HOLD TIME)		
					9380-P-Perchlorate		
					9052A-Nitrite (8 HOUR HOLD TIME)		
					9010C-9020A/470A-Dissolved Metals		
					9010C/9020A/470A-Total Metals		
					9020A-Aminec		
					9020B-Amine		
					9020C-Phosphorous		
					9020D-Nitrate (8 HOUR HOLD TIME)		
					9020E-Nitrite (8 HOUR HOLD TIME)		
					9020F-Sulfide (8 HOUR HOLD TIME)		
					9020G-Methane		
					9020H-Methanol		
					9020I-Ethanol		
					9020J-Glycol		
					9020K-Other (specify)		
					Other:		
					Special Instructions/Note:		
					QC dep		
					Filtered		
					Filtered		
					QC dep		

Possibly Hazardous:  Non-Hazard  Flammable  Skin Irritant  Poison A  Unknown  Radiological  
Deliverable Requested: I, II, III, IV, Other (specify)

Sample Disposal / A fee may be assessed if samples are retained longer than 1 month)  
 Return To Client  Disposal By Lab  Archive For \_\_\_\_\_ months

Special Instructions/OC Requirements:  
 Method of Shipment: 2

Empty Kit Relinquished by:	Date/Time:	Received by:
<u>Relinquished by</u>	6-20-18/1600	Company Cardno
Relinquished by:	Date/Time: 6/20/18 - 1600	Received by: <u>Leed</u>
Relinquished by:	Date/Time:	Received by:
Custody Seal intact: Yes <input checked="" type="checkbox"/> No <input type="checkbox"/>	Custody Seal No.:	Cooler Temperature(s) °C and Other Remarks:

## Login Sample Receipt Checklist

Client: Cardno GS, Inc

Job Number: 280-111519-2

**Login Number: 111519**

**List Source: TestAmerica Denver**

**List Number: 1**

**Creator: Pottruff, Reed W**

Question	Answer	Comment
Radioactivity wasn't checked or is </= background as measured by a survey meter.	N/A	
The cooler's custody seal, if present, is intact.	True	
Sample custody seals, if present, are intact.	True	
The cooler or samples do not appear to have been compromised or tampered with.	True	
Samples were received on ice.	True	
Cooler Temperature is acceptable.	True	
Cooler Temperature is recorded.	True	
COC is present.	True	
COC is filled out in ink and legible.	True	
COC is filled out with all pertinent information.	True	
Is the Field Sampler's name present on COC?	True	
There are no discrepancies between the containers received and the COC.	True	
Samples are received within Holding Time (excluding tests with immediate HTs)	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	False	Refer to Job Narrative for details.
Sample collection date/times are provided.	True	
Appropriate sample containers are used.	True	
Sample bottles are completely filled.	True	
Sample Preservation Verified.	True	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	True	
Containers requiring zero headspace have no headspace or bubble is <6mm (1/4").	False	Refer to Job Narrative for details.
Multiphasic samples are not present.	True	
Samples do not require splitting or compositing.	True	
Residual Chlorine Checked.	N/A	